

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122023\
 Data File : VY016747.D
 Acq On : 21 Dec 2023 07:54
 Operator : SY/MD
 Sample : 05876-16
 Misc : 4.69g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COLY1

Quant Time: Dec 21 08:14:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 21 01:49:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.825	168	113099	50.000	ug/l	0.02
34) 1,4-Difluorobenzene	8.721	114	339878	50.000	ug/l	0.02
63) Chlorobenzene-d5	11.513	117	380735	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.446	152	153834	50.000	ug/l	0.01

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.173	65	98563	95.463	ug/l	0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	=	190.920%#
35) Dibromofluoromethane	7.752	113	120589	61.939	ug/l	0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	=	123.880%
50) Toluene-d8	10.209	98	447473	63.757	ug/l	0.02
Spiked Amount	50.000	Range	58 - 134	Recovery	=	127.520%
62) 4-Bromofluorobenzene	12.507	95	163039	71.603	ug/l	0.02
Spiked Amount	50.000	Range	30 - 143	Recovery	=	143.200%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122023\
Data File : VY016747.D
Acq On : 21 Dec 2023 07:54
Operator : SY/MD
Sample : 05876-16
Misc : 4.69g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COLY1

Quant Time: Dec 21 08:14:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 21 01:49:45 2023
Response via : Initial Calibration

