

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122023\
 Data File : VY016750.D
 Acq On : 21 Dec 2023 09:04
 Operator : SY/MD
 Sample : 05876-19
 Misc : 5.40g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COLY4

Quant Time: Dec 21 14:20:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 21 01:49:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.813	168	124091	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.709	114	386708	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.501	117	481218	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	243606	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.167	65	117582	103.796	ug/l	0.01
Spiked Amount	50.000	Range 50 - 163	Recovery	=	207.600%	#
35) Dibromofluoromethane	7.740	113	142002	64.105	ug/l	0.01
Spiked Amount	50.000	Range 54 - 147	Recovery	=	128.200%	
50) Toluene-d8	10.197	98	521299	65.281	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	=	130.560%	
62) 4-Bromofluorobenzene	12.495	95	237460	91.658	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	=	183.320%	#
Target Compounds						
20) Methylene Chloride	4.753	84	9610	3.343	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122023\
Data File : VY016750.D
Acq On : 21 Dec 2023 09:04
Operator : SY/MD
Sample : 05876-19
Misc : 5.40g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COLY4

Quant Time: Dec 21 14:20:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 21 01:49:45 2023
Response via : Initial Calibration

