



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122023\
Data File : VY016723.D
Acq On : 20 Dec 2023 13:56
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
Title : SW846 8260
Last Update : Thu Dec 21 01:49:45 2023



AutoFind: Scans 1750, 1751, 1752; Background Corrected with Scan 1741

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.1	12390	PASS
75	95	30	60	52.0	33813	PASS
95	95	100	100	100.0	65008	PASS
96	95	5	9	6.7	4333	PASS
173	174	0.00	2	1.2	624	PASS
174	95	50	100	81.3	52837	PASS
175	174	5	9	7.4	3936	PASS
176	174	95	101	95.5	50475	PASS
177	176	5	9	6.8	3426	PASS