

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122122\
 Data File : VY011951.D
 Acq On : 21 Dec 2022 16:57
 Operator : KP/MD
 Sample : N6088-01
 Misc : 4.84g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B-27-5B01

Quant Time: Dec 22 00:14:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 20 12:37:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.789	168	194138	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	343594	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	316817	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	114239	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	119778	61.239	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 122.480%	
35) Dibromofluoromethane	7.716	113	95393	46.116	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 92.240%	
50) Toluene-d8	10.179	98	417519	49.749	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 99.500%	
62) 4-Bromofluorobenzene	12.477	95	130354	46.182	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 92.360%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.906	85	114	2.537	ug/l #	41
11) Tert butyl alcohol	4.954	59	471	Below Cal	#	74
16) Acetone	3.942	43	19058	36.875	ug/l	95
17) Carbon Disulfide	4.192	76	26606	4.541	ug/l	98
25) 2-Butanone	6.990	43	4284	6.575	ug/l	96
31) Cyclohexane	7.783	56	5467	1.660	ug/l #	37
36) 1,1-Dichloropropene	7.783	75	13850	4.174	ug/l #	50
37) Ethyl Acetate	6.990	43	4284	2.878	ug/l #	80
38) Carbon Tetrachloride	7.789	117	19776	5.161	ug/l #	18
39) Methylcyclohexane	9.179	83	7356	1.842	ug/l	96
40) Benzene	8.161	78	13152	1.385	ug/l	95
52) Toluene	10.246	92	6858	1.134	ug/l	98
55) 1,1,2-Trichloroethane	10.618	97	16801	9.369	ug/l #	21
56) Ethyl methacrylate	10.520	69	2738	1.206	ug/l #	1
59) 2-Hexanone	10.801	43	14389	13.492	ug/l #	29
68) m/p-Xylenes	11.697	106	6300	1.362	ug/l	74
69) o-Xylene	12.026	106	6623	1.539	ug/l	68
73) Isopropylbenzene	12.331	105	13731	1.620	ug/l	75
74) N-amyl acetate	12.123	43	77081	44.352	ug/l #	64
75) 1,1,2,2-Tetrachloroethane	12.526	83	10647	7.529	ug/l #	20
76) 1,2,3-Trichloropropane	12.483	75	67298	47.342	ug/l #	100
79) 2-Chlorotoluene	12.812	91	7477	1.305	ug/l	72
80) 1,3,5-Trimethylbenzene	12.812	105	40610	5.595	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.483	75	67298	129.647	ug/l #	13
82) 4-Chlorotoluene	12.812	91	6987	1.179	ug/l	78
84) 1,2,4-Trimethylbenzene	13.117	105	45863	6.423	ug/l	96
86) p-Isopropyltoluene	13.367	119	23222	2.996	ug/l	96
90) Hexachloroethane	13.971	117	2155	1.457	ug/l #	9
92) 1,2-Dibromo-3-Chloropr...	14.336	75	257	1.057	ug/l #	1
95) Naphthalene	15.233	128	115498	28.705	ug/l #	89
96) 1,2,3-Trichlorobenzene	15.391	180	4719	2.759	ug/l #	58


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122122\
Data File : VY011951.D
Acq On : 21 Dec 2022 16:57
Operator : KP/MD
Sample : N6088-01
Misc : 4.84g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-27-5B01

Quant Time: Dec 22 00:14:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 20 12:37:37 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57




Instrument :
MSVOA_Y
ClientSampleId :
B-27-5B01


Tgt Ion:168 Resp: 194138
Ion Ratio Lower Upper
168 100
99 57.0 44.0 66.0



#2
Dichlorodifluoromethane
Concen: 2.537 ug/l
RT: 1.906 min Scan# 14
Delta R.T. -0.000 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57

Tgt Ion: 85 Resp: 114
Ion Ratio Lower Upper
85 100
87 0.0 16.8 50.4#









#11
Tert butyl alcohol
Concen: Below Cal
RT: 4.954 min Scan# 513
Delta R.T. 0.006 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57




Instrument :
MSVOA_Y
ClientSampleId :
B-27-5B01


Tgt Ion: 59 Resp: 471
Ion Ratio Lower Upper
59 100
57 0.0 7.4 11.2#



#16
Acetone
Concen: 36.875 ug/l
RT: 3.942 min Scan# 348
Delta R.T. -0.006 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57

Tgt Ion: 43 Resp: 19058
Ion Ratio Lower Upper
43 100
58 32.2 27.9 41.9









#17
Carbon Disulfide
Concen: 4.541 ug/l
RT: 4.192 min Scan# 31
Delta R.T. -0.006 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57




Instrument :
MSVOA_Y
ClientSampleId :
B-27-5B01


Tgt Ion: 76 Resp: 26606
Ion Ratio Lower Upper
76 100
78 9.9 7.2 10.8



#25
2-Butanone
Concen: 6.575 ug/l
RT: 6.990 min Scan# 848
Delta R.T. 0.006 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57


Tgt Ion: 43 Resp: 4284
Ion Ratio Lower Upper
43 100
72 30.4 22.6 34.0








#31
Cyclohexane
Concen: 1.660 ug/l
RT: 7.783 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57

Instrument :
MSVOA_Y
ClientSampleId :
B-27-5B01







Tgt Ion: 56 Resp: 5467
Ion Ratio Lower Upper
56 100
69 122.5 29.2 43.8#
84 119.5 72.6 108.8#



#33
1,2-Dichloroethane-d4
Concen: 61.239 ug/l
RT: 8.143 min Scan# 1037
Delta R.T. -0.001 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57

Tgt Ion: 65 Resp: 119778
Ion Ratio Lower Upper
65 100
67 51.6 0.0 103.2









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.691 min Scan# 1127
Delta R.T. 0.000 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57

Instrument :
MSVOA_Y
ClientSampleId :
B-27-5B01

Tgt Ion:114 Resp: 343594




Ion	Ratio	Lower	Upper
114	100		
63	18.4	0.0	39.6
88	14.9	0.0	32.2




#35
Dibromofluoromethane
Concen: 46.116 ug/l
RT: 7.716 min Scan# 967
Delta R.T. -0.000 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57

Tgt Ion:113 Resp: 95393

Ion	Ratio	Lower	Upper
113	100		
111	103.6	83.0	124.4
192	21.2	16.0	24.0









#36
1,1-Dichloropropene
Concen: 4.174 ug/l
RT: 7.783 min Scan# 91
Delta R.T. -0.134 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57




Instrument :
MSVOA_Y
ClientSampleId :
B-27-5B01


Tgt Ion: 75 Resp: 13850
Ion Ratio Lower Upper
75 100
110 9.6 18.1 54.1#
77 0.0 25.1 37.7#



#37
Ethyl Acetate
Concen: 2.878 ug/l
RT: 6.990 min Scan# 848
Delta R.T. -0.080 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57

Tgt Ion: 43 Resp: 4284
Ion Ratio Lower Upper
43 100
61 8.0 11.3 16.9#
70 0.0 8.2 12.4#









#38
Carbon Tetrachloride
Concen: 5.161 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.116 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57




Instrument :
MSVOA_Y
ClientSampleId :
B-27-5B01


Tgt Ion:117 Resp: 19776
Ion Ratio Lower Upper
117 100
119 5.6 73.8 110.8#
121 0.0 25.2 37.8#



#39
Methylcyclohexane
Concen: 1.842 ug/l
RT: 9.179 min Scan# 1207
Delta R.T. -0.006 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57

Tgt Ion: 83 Resp: 7356
Ion Ratio Lower Upper
83 100
55 77.4 58.2 87.2
98 49.5 38.6 58.0









#40
Benzene
Concen: 1.385 ug/l
RT: 8.161 min Scan# 1040
Delta R.T. -0.000 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57




Instrument :
MSVOA_Y
ClientSampleId :
B-27-5B01


Tgt Ion: 78 Resp: 13152
Ion Ratio Lower Upper
78 100
77 25.5 18.6 27.8



#50
Toluene-d8
Concen: 49.749 ug/l
RT: 10.179 min Scan# 1371
Delta R.T. -0.000 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57

Tgt Ion: 98 Resp: 417519
Ion Ratio Lower Upper
98 100
100 63.9 51.9 77.9









#52
Toluene
Concen: 1.134 ug/l
RT: 10.246 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57




Instrument :
MSVOA_Y
ClientSampleId :
B-27-5B01


Tgt Ion: 92 Resp: 6858
Ion Ratio Lower Upper
92 100
91 174.7 137.4 206.0



#55
1,1,2-Trichloroethane
Concen: 9.369 ug/l
RT: 10.618 min Scan# 1443
Delta R.T. -0.031 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57

Tgt Ion: 97 Resp: 16801
Ion Ratio Lower Upper
97 100
83 10.5 67.7 101.5#
85 1.6 43.8 65.8#
99 0.0 49.0 73.4#









#56
Ethyl methacrylate
Concen: 1.206 ug/l
RT: 10.520 min Scan# 1427
Delta R.T. 0.012 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57




Instrument :
MSVOA_Y
ClientSampleId :
B-27-5B01


Tgt Ion: 69 Resp: 2738
Ion Ratio Lower Upper
69 100
41 159.1 48.9 73.3#
39 84.2 25.9 38.9#



#59
2-Hexanone
Concen: 13.492 ug/l
RT: 10.801 min Scan# 1473
Delta R.T. -0.031 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57

Tgt Ion: 43 Resp: 14389
Ion Ratio Lower Upper
43 100
58 4.4 28.4 85.2#









#62
4-Bromofluorobenzene
Concen: 46.182 ug/l
RT: 12.477 min Scan# 11
Delta R.T. -0.006 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57




Instrument :
MSVOA_Y
ClientSampleId :
B-27-5B01


Tgt Ion: 95 Resp: 130354
Ion Ratio Lower Upper
95 100
174 84.5 0.0 172.0
176 82.1 0.0 167.6



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.489 min Scan# 1586
Delta R.T. -0.000 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57

Tgt Ion: 117 Resp: 316817
Ion Ratio Lower Upper
117 100
82 54.2 44.8 67.2
119 31.9 25.4 38.2









#68
m/p-Xylenes
Concen: 1.362 ug/l
RT: 11.697 min Scan# 1620
Delta R.T. -0.006 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57




Instrument :
MSVOA_Y
ClientSampleId :
B-27-5B01


Tgt Ion:106 Resp: 6300
Ion Ratio Lower Upper
106 100
91 242.9 162.5 243.7



#69
o-Xylene
Concen: 1.539 ug/l
RT: 12.026 min Scan# 1674
Delta R.T. -0.006 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57

Tgt Ion:106 Resp: 6623
Ion Ratio Lower Upper
106 100
91 266.2 107.5 322.4









#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. -0.000 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57




Instrument :
MSVOA_Y
ClientSampleId :
B-27-5B01


Tgt Ion:152 Resp: 114239
Ion Ratio Lower Upper
152 100
115 59.8 29.1 87.3
150 155.9 0.0 342.6



#73
Isopropylbenzene
Concen: 1.620 ug/l
RT: 12.331 min Scan# 1724
Delta R.T. -0.000 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57

Tgt Ion:105 Resp: 13731
Ion Ratio Lower Upper
105 100
120 13.6 13.2 39.6









#74
N-amy1 acetate
Concen: 44.352 ug/l
RT: 12.123 min Scan# 1
Delta R.T. -0.018 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57

Instrument :
MSVOA_Y
ClientSampleId :
B-27-5B01

Tgt Ion: 43 Resp: 77081




Ion	Ratio	Lower	Upper
43	100		
70	19.5	35.8	53.8#
55	16.9	21.0	31.4#
61	0.0	20.0	30.0#




#75
1,1,2,2-Tetrachloroethane
Concen: 7.529 ug/l
RT: 12.526 min Scan# 1756
Delta R.T. -0.055 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57

Tgt Ion: 83 Resp: 10647

Ion	Ratio	Lower	Upper
83	100		
131	0.0	5.6	16.8#
85	134.2	32.3	96.8#









#76
1,2,3-Trichloropropane
Concen: 47.342 ug/l
RT: 12.483 min Scan# 11
Delta R.T. -0.153 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57




Instrument :
MSVOA_Y
ClientSampleId :
B-27-5B01


Tgt Ion: 75 Resp: 67298
Ion Ratio Lower Upper
75 100
77 1.8 0.0 0.0#



#79
2-Chlorotoluene
Concen: 1.305 ug/l
RT: 12.812 min Scan# 1803
Delta R.T. 0.055 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57

Tgt Ion: 91 Resp: 7477
Ion Ratio Lower Upper
91 100
126 18.2 17.2 51.5









#80
1,3,5-Trimethylbenzene
Concen: 5.595 ug/l
RT: 12.812 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57




Instrument :
MSVOA_Y
ClientSampleId :
B-27-5B01


Tgt Ion:105 Resp: 40610
Ion Ratio Lower Upper
105 100
120 49.2 24.2 72.6



#81
trans-1,4-Dichloro-2-butene
Concen: 129.647 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.104 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57

Tgt Ion: 75 Resp: 67298
Ion Ratio Lower Upper
75 100
53 0.0 72.5 108.7#
89 0.0 37.0 55.4#









#82
4-Chlorotoluene
Concen: 1.179 ug/l
RT: 12.812 min Scan# 1803
Delta R.T. -0.043 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57




Instrument :
MSVOA_Y
ClientSampleId :
B-27-5B01


Tgt Ion: 91 Resp: 6987
Ion Ratio Lower Upper
91 100
126 21.6 17.1 51.2



#84
1,2,4-Trimethylbenzene
Concen: 6.423 ug/l
RT: 13.117 min Scan# 1853
Delta R.T. -0.000 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57

Tgt Ion: 105 Resp: 45863
Ion Ratio Lower Upper
105 100
120 42.9 22.9 68.7





#86

p-Isopropyltoluene

Concen: 2.996 ug/l

RT: 13.367 min Scan# 119

Delta R.T. -0.000 min

Lab File: VY011951.D

Acq: 21 Dec 2022 16:57

Instrument :

MSVOA_Y

ClientSampleId :

B-27-5B01





Tgt Ion:119 Resp: 23222

Ion Ratio Lower Upper

119 100

134 24.3 13.0 39.0

91 22.0 11.9 35.9



#90

Hexachloroethane

Concen: 1.457 ug/l

RT: 13.971 min Scan# 1993

Delta R.T. 0.012 min

Lab File: VY011951.D




Acq: 21 Dec 2022 16:57


Tgt Ion:117 Resp: 2155

Ion Ratio Lower Upper

117 100

201 0.0 40.3 120.8#









#92
1,2-Dibromo-3-Chloropropane
Concen: 1.057 ug/l
RT: 14.336 min Scan# 2056
Delta R.T. -0.018 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57




Instrument :
MSVOA_Y
ClientSampleId :
B-27-5B01


Tgt Ion: 75 Resp: 257
Ion Ratio Lower Upper
75 100
155 1519.5 42.3 126.8#
157 0.0 54.4 163.3#



#95
Naphthalene
Concen: 28.705 ug/l
RT: 15.233 min Scan# 2200
Delta R.T. -0.006 min
Lab File: VY011951.D
Acq: 21 Dec 2022 16:57

Tgt Ion: 128 Resp: 115498
Ion Ratio Lower Upper
128 100
127 20.4 10.2 15.2#
129 10.8 8.8 13.2





#96
 1,2,3-Trichlorobenzene
 Concen: 2.759 ug/l
 RT: 15.391 min Scan# 21
 Delta R.T. -0.031 min
 Lab File: VY011951.D
 Acq: 21 Dec 2022 16:57

Instrument :
 MSVOA_Y
 ClientSampleId :
 B-27-5B01

Tgt Ion:180 Resp: 4719
 Ion Ratio Lower Upper
 180 100
 182 41.9 47.9 143.7#
 145 33.7 16.1 48.2

