

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122122\
 Data File : VY011946.D
 Acq On : 21 Dec 2022 15:00
 Operator : KP/MD
 Sample : N6098-11
 Misc : 5.86g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 FVSUB-TP2-COMP-02

Quant Time: Dec 22 00:13:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 20 12:37:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	307	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	646	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	252	50.000	ug/l	# 0.01
72) 1,4-Dichlorobenzene-d4	13.434	152	20	50.000	ug/l	# 0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	269	86.972	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 173.940%#	
35) Dibromofluoromethane	7.716	113	175	44.997	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 90.000%	
50) Toluene-d8	10.185	98	437	27.695	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 55.400%	
62) 4-Bromofluorobenzene	12.483	95	41	7.726	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 15.460%#	
Target Compounds						
10) Methyl Iodide	4.107	142	40	9.956	ug/l	# 38

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122122\
Data File : VY011946.D
Acq On : 21 Dec 2022 15:00
Operator : KP/MD
Sample : N6098-11
Misc : 5.86g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
FVSUB-TP2-COMP-02

Quant Time: Dec 22 00:13:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 20 12:37:37 2022
Response via : Initial Calibration

