

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122122\
 Data File : VY011948.D
 Acq On : 21 Dec 2022 15:47
 Operator : KP/MD
 Sample : N6092-02
 Misc : 4.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 20221259-LOCATION-7

Quant Time: Dec 22 00:14:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 20 12:37:37 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	204398	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	355253	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	313268	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	120067	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	118379	57.486	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 114.980%		
35) Dibromofluoromethane	7.716	113	110033	51.448	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 102.900%		
50) Toluene-d8	10.179	98	428036	49.328	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 98.660%		
62) 4-Bromofluorobenzene	12.483	95	120755	41.378	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 82.760%		
Target Compounds						Qvalue
16) Acetone	3.942	43	55756	93.795	ug/l	96
17) Carbon Disulfide	4.192	76	14905	2.416	ug/l	99
25) 2-Butanone	6.984	43	15058	21.952	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122122\
Data File : VY011948.D
Acq On : 21 Dec 2022 15:47
Operator : KP/MD
Sample : N6092-02
Misc : 4.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
20221259-LOCATION-7

Quant Time: Dec 22 00:14:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 20 12:37:37 2022
Response via : Initial Calibration

