

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122122\
 Data File : VY011957.D
 Acq On : 21 Dec 2022 19:18
 Operator : KP/MD
 Sample : N6129-04
 Misc : 5.29g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 D8

Quant Time: Dec 22 00:15:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 20 12:37:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	209209	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	367704	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	346603	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	154243	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	126695	60.110	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	120.220%
35) Dibromofluoromethane	7.716	113	115253	52.064	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.120%
50) Toluene-d8	10.179	98	447782	49.856	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.720%
62) 4-Bromofluorobenzene	12.483	95	142145	47.058	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	94.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122122\
Data File : VY011957.D
Acq On : 21 Dec 2022 19:18
Operator : KP/MD
Sample : N6129-04
Misc : 5.29g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
D8

Quant Time: Dec 22 00:15:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 20 12:37:37 2022
Response via : Initial Calibration

