

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122122\
 Data File : VY011960.D
 Acq On : 21 Dec 2022 20:28
 Operator : KP/MD
 Sample : N6161-08
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 N40305

Quant Time: Dec 22 00:15:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 20 12:37:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	19579	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	30019	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	24325	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	9284	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	10465	53.053	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.100%
35) Dibromofluoromethane	7.722	113	9661	53.458	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.920%
50) Toluene-d8	10.179	98	35106	47.878	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.760%
62) 4-Bromofluorobenzene	12.483	95	8855	35.908	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	71.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122122\
Data File : VY011960.D
Acq On : 21 Dec 2022 20:28
Operator : KP/MD
Sample : N6161-08
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
N40305

Quant Time: Dec 22 00:15:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 20 12:37:37 2022
Response via : Initial Calibration

