

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122623\
 Data File : VY016851.D
 Acq On : 26 Dec 2023 14:32
 Operator : SY/MD
 Sample : 06018-01
 Misc : 5.27g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMP-1

Quant Time: Dec 27 01:34:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 21 01:49:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	106055	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	230056	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	235445	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	93325	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	67784	70.013	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	140.020%
35) Dibromofluoromethane	7.734	113	67614	51.308	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.620%
50) Toluene-d8	10.191	98	244924	51.557	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	103.120%
62) 4-Bromofluorobenzene	12.489	95	83013	53.861	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	107.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122623\
Data File : VY016851.D
Acq On : 26 Dec 2023 14:32
Operator : SY/MD
Sample : 06018-01
Misc : 5.27g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMP-1

Quant Time: Dec 27 01:34:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 21 01:49:45 2023
Response via : Initial Calibration

