

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122623\
 Data File : VY016855.D
 Acq On : 26 Dec 2023 16:05
 Operator : SY/MD
 Sample : 06015-05
 Misc : 6.83g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MR-323-0-2

Quant Time: Dec 27 01:35:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 21 01:49:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	104909	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	229145	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	232361	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	89490	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	65501	68.394	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 136.780%	
35) Dibromofluoromethane	7.734	113	66517	50.676	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 101.360%	
50) Toluene-d8	10.191	98	239393	50.593	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 101.180%	
62) 4-Bromofluorobenzene	12.489	95	79437	51.746	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 103.500%	
Target Compounds						
					Qvalue	
30) Chloroform	7.514	83	2361	1.119	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122623\
Data File : VY016855.D
Acq On : 26 Dec 2023 16:05
Operator : SY/MD
Sample : 06015-05
Misc : 6.83g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MR-323-0-2

Quant Time: Dec 27 01:35:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 21 01:49:45 2023
Response via : Initial Calibration

