

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122623\
 Data File : VY016857.D
 Acq On : 26 Dec 2023 16:52
 Operator : SY/MD
 Sample : 06015-01
 Misc : 5.76g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MR-317-0-2

Quant Time: Dec 27 01:36:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 21 01:49:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	102530	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	222859	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	226781	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	85317	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	62120	66.368	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	132.740%
35) Dibromofluoromethane	7.728	113	64413	50.457	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.920%
50) Toluene-d8	10.191	98	231352	50.272	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.540%
62) 4-Bromofluorobenzene	12.489	95	75760	50.743	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	101.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122623\
Data File : VY016857.D
Acq On : 26 Dec 2023 16:52
Operator : SY/MD
Sample : 06015-01
Misc : 5.76g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MR-317-0-2

Quant Time: Dec 27 01:36:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 21 01:49:45 2023
Response via : Initial Calibration

