

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122922\
 Data File : VY012034.D
 Acq On : 29 Dec 2022 17:35
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 30 04:01:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 20 12:37:37 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	80	0.00
2 T	Dichlorodifluoromethane	0.377	0.399	-5.8	72	0.00
3 P	Chloromethane	0.445	0.384	13.7	64	0.00
4 C	Vinyl Chloride	0.491	0.444	9.6#	67	0.00
5 T	Bromomethane	0.349	0.317	9.2	71	0.00
6 T	Chloroethane	0.321	0.298	7.2	70	0.00
7 T	Trichlorofluoromethane	0.859	0.853	0.7	75	0.00
8 T	Diethyl Ether	0.265	0.284	-7.2	82	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.473	0.466	1.5	76	0.00
10 T	Methyl Iodide	0.654	0.624	4.6	68	0.00
11 T	Tert butyl alcohol	0.065	0.055	15.4	89	0.00
12 CM	1,1-Dichloroethene	0.473	0.447	5.5#	72	0.00
13 T	Acrolein	0.047	0.040	14.9	73	0.00
14 T	Allyl chloride	0.697	0.643	7.7	70	0.00
15 T	Acrylonitrile	0.123	0.146	-18.7	89	0.00
16 T	Acetone	0.137	0.134	2.2	65	0.00
17 T	Carbon Disulfide	1.509	1.201	20.4	60	0.00
18 T	Methyl Acetate	0.348	0.390	-12.1	87	0.00
19 T	Methyl tert-butyl Ether	1.214	1.451	-19.5	89	0.00
20 T	Methylene Chloride	0.701	0.558	20.4	73	0.00
21 T	trans-1,2-Dichloroethene	0.538	0.517	3.9	72	0.00
22 T	Diisopropyl ether	1.372	1.412	-2.9	77	0.00
23 T	Vinyl Acetate	0.789	0.883	-11.9	81	0.00
24 P	1,1-Dichloroethane	0.892	0.871	2.4	76	0.00
25 T	2-Butanone	0.168	0.180	-7.1	76	0.00
26 T	2,2-Dichloropropane	0.873	0.871	0.2	78	0.00
27 T	cis-1,2-Dichloroethene	0.581	0.595	-2.4	78	0.00
28 T	Bromochloromethane	0.245	0.228	6.9	76	0.00
29 T	Tetrahydrofuran	0.095	0.112	-17.9	85	0.00
30 C	Chloroform	0.961	1.012	-5.3#	81	0.00
31 T	Cyclohexane	0.848	0.718	15.3	66	0.00
32 T	1,1,1-Trichloroethane	0.914	0.939	-2.7	80	0.00
33 S	1,2-Dichloroethane-d4	0.504	0.555	-10.1	87	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	80	0.00
35 S	Dibromofluoromethane	0.301	0.321	-6.6	83	0.00
36 T	1,1-Dichloropropene	0.483	0.448	7.2	71	0.00
37 T	Ethyl Acetate	0.217	0.250	-15.2	89	0.00
38 T	Carbon Tetrachloride	0.558	0.566	-1.4	78	0.00
39 T	Methylcyclohexane	0.581	0.538	7.4	69	0.00
40 TM	Benzene	1.382	1.340	3.0	75	0.00
41 T	Methacrylonitrile	0.105	0.108	-2.9	74	0.00
42 TM	1,2-Dichloroethane	0.398	0.439	-10.3	86	0.00
43 T	Isopropyl Acetate	0.396	0.461	-16.4	87	0.00
44 TM	Trichloroethene	0.389	0.393	-1.0	79	0.00
45 C	1,2-Dichloropropane	0.319	0.314	1.6#	77	0.00
46 T	Dibromomethane	0.184	0.203	-10.3	85	0.00
47 T	Bromodichloromethane	0.467	0.508	-8.8	83	0.00
48 T	Methyl methacrylate	0.183	0.210	-14.8	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122922\
 Data File : VY012034.D
 Acq On : 29 Dec 2022 17:35
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 30 04:01:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 20 12:37:37 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.003	-50.0#	96	0.00
50 S	Toluene-d8	1.221	1.214	0.6	77	0.00
51 T	4-Methyl-2-Pentanone	0.211	0.256	-21.3	89	0.00
52 CM	Toluene	0.880	0.897	-1.9#	78	0.00
53 T	t-1,3-Dichloropropene	0.471	0.530	-12.5	85	0.00
54 T	cis-1,3-Dichloropropene	0.530	0.566	-6.8	81	0.00
55 T	1,1,2-Trichloroethane	0.261	0.293	-12.3	88	0.00
56 T	Ethyl methacrylate	0.330	0.408	-23.6	90	0.00
57 T	1,3-Dichloropropane	0.436	0.489	-12.2	84	0.00
58 T	2-Chloroethyl Vinyl ether	0.127	0.142	-11.8	89	0.00
59 T	2-Hexanone	0.155	0.184	-18.7	83	0.00
60 T	Dibromochloromethane	0.329	0.380	-15.5	87	0.00
61 T	1,2-Dibromoethane	0.244	0.276	-13.1	86	0.00
62 S	4-Bromofluorobenzene	0.411	0.448	-9.0	85	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	83	0.00
64 T	Tetrachloroethene	0.438	0.451	-3.0	82	0.00
65 PM	Chlorobenzene	1.038	1.039	-0.1	80	0.00
66 T	1,1,1,2-Tetrachloroethane	0.386	0.408	-5.7	84	0.00
67 C	Ethyl Benzene	1.877	1.876	0.1#	78	0.00
68 T	m/p-Xylenes	0.730	0.739	-1.2	79	0.00
69 T	o-Xylene	0.679	0.707	-4.1	81	0.00
70 T	Styrene	1.140	1.220	-7.0	83	0.00
71 P	Bromoform	0.229	0.272	-18.8	94	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
73 T	Isopropylbenzene	3.710	3.610	2.7	81	0.00
74 T	N-amyl acetate	0.761	0.847	-11.3	90	0.00
75 P	1,1,2,2-Tetrachloroethane	0.619	0.652	-5.3	90	0.00
76 T	1,2,3-Trichloropropane	0.622	0.532	14.5	72	0.00
77 T	Bromobenzene	0.858	0.864	-0.7	86	0.00
78 T	n-propylbenzene	4.460	4.275	4.1	80	0.00
79 T	2-Chlorotoluene	2.507	2.422	3.4	82	0.00
80 T	1,3,5-Trimethylbenzene	3.177	3.121	1.8	82	0.00
81 T	trans-1,4-Dichloro-2-butene	0.227	0.248	-9.3	91	0.00
82 T	4-Chlorotoluene	2.594	2.530	2.5	82	0.00
83 T	tert-Butylbenzene	2.692	2.680	0.4	82	0.00
84 T	1,2,4-Trimethylbenzene	3.125	3.122	0.1	83	0.00
85 T	sec-Butylbenzene	4.020	3.896	3.1	81	0.00
86 T	p-Isopropyltoluene	3.392	3.387	0.1	82	0.00
87 T	1,3-Dichlorobenzene	1.726	1.696	1.7	84	0.00
88 T	1,4-Dichlorobenzene	1.716	1.699	1.0	85	0.00
89 T	n-Butylbenzene	3.071	2.997	2.4	81	0.00
90 T	Hexachloroethane	0.647	0.624	3.6	82	0.00
91 T	1,2-Dichlorobenzene	1.505	1.553	-3.2	88	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.106	0.130	-22.6	101	0.00
93 T	1,2,4-Trichlorobenzene	0.885	1.030	-16.4	94	0.00
94 T	Hexachlorobutadiene	0.572	0.600	-4.9	89	0.00
95 T	Naphthalene	1.644	2.223	-35.2#	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122922\
Data File : VY012034.D
Acq On : 29 Dec 2022 17:35
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Dec 30 04:01:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 20 12:37:37 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.749	0.915	-22.2	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6