

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD010220\
 Data File : VD064714.D
 Acq On : 02 Jan 2020 13:36
 Operator : VA/SY
 Sample : K6464-01
 Misc : 6.01G/5.00ml/MSVOA D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-4-(5-6)

Quant Time: Jan 03 00:59:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 24 13:19:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	501715	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	765879	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	662796	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	292584	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	236839	53.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.06%	
35) Dibromofluoromethane	7.91	113	240839	51.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.52%	
50) Toluene-d8	10.34	98	906569	50.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.76%	
62) 4-Bromofluorobenzene	12.64	95	275905	48.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.98%	
Target Compounds						
16) Acetone	4.17	43	21213	24.351	ug/l	98
17) Carbon Disulfide	4.40	76	16336	1.145	ug/l #	93
30) Chloroform	7.71	83	11912	1.378	ug/l	99
86) p-Isopropyltoluene	13.53	119	24540	1.255	ug/l	96
95) Naphthalene	15.41	128	16910	1.738	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD010220\
Data File : VD064714.D
Acq On : 02 Jan 2020 13:36
Operator : VA/SY
Sample : K6464-01
Misc : 6.01G/5.00ml/MSVOA D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-4-(5-6)

Quant Time: Jan 03 00:59:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 24 13:19:06 2019
Response via : Initial Calibration

