

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD010220\
 Data File : VD064715.D
 Acq On : 02 Jan 2020 14:05
 Operator : VA/SY
 Sample : K6464-02
 Misc : 5.08G/5.00ml/MSVOA D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-5-(8-10)

Quant Time: Jan 03 01:01:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 24 13:19:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	279341	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	430403	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	375150	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	156091	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	166896	67.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	135.52%	
35) Dibromofluoromethane	7.92	113	144931	54.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.78%	
50) Toluene-d8	10.34	98	510274	50.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.92%	
62) 4-Bromofluorobenzene	12.64	95	150790	47.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.30%	
Target Compounds						
16) Acetone	4.17	43	27002	50.821	ug/l	97
30) Chloroform	7.71	83	6379	1.325	ug/l	88
95) Naphthalene	15.42	128	10805	2.082	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD010220\
Data File : VD064715.D
Acq On : 02 Jan 2020 14:05
Operator : VA/SY
Sample : K6464-02
Misc : 5.08G/5.00ml/MSVOA D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-5-(8-10)

Quant Time: Jan 03 01:01:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 24 13:19:06 2019
Response via : Initial Calibration

