

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD010220\
 Data File : VD064722.D
 Acq On : 02 Jan 2020 17:26
 Operator : VA/SY
 Sample : K6465-05
 Misc : 6.40G/5.00ml/MSVOA D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 GP-9-(1-3)

Quant Time: Jan 03 01:37:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 24 13:19:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	497667	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	762511	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	685906	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	314344	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.34	65	235169	53.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.18%	
35) Dibromofluoromethane	7.92	113	238601	51.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.00%	
50) Toluene-d8	10.34	98	920544	51.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.76%	
62) 4-Bromofluorobenzene	12.64	95	277259	48.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.88%	
Target Compounds						
30) Chloroform	7.72	83	10535	1.228	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD010220\
Data File : VD064722.D
Acq On : 02 Jan 2020 17:26
Operator : VA/SY
Sample : K6465-05
Misc : 6.40G/5.00ml/MSVOA D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
GP-9-(1-3)

Quant Time: Jan 03 01:37:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 24 13:19:06 2019
Response via : Initial Calibration

