

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD010224\
 Data File : VD078115.D
 Acq On : 02 Jan 2024 09:23
 Operator : RP/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 03 00:02:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122623S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 27 03:40:16 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	111	0.00
2 T	Dichlorodifluoromethane	0.429	0.387	9.8	111	0.00
3 P	Chloromethane	0.591	0.510	13.7	97	0.00
4 C	Vinyl Chloride	0.891	0.791	11.2#	97	0.00
5 T	Bromomethane	0.714	0.619	13.3	103	0.00
6 T	Chloroethane	0.655	0.579	11.6	99	0.00
7 T	Trichlorofluoromethane	0.857	0.867	-1.2	116	0.00
8 T	Diethyl Ether	0.235	0.232	1.3	115	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.534	0.549	-2.8	119	0.00
10 T	Methyl Iodide	0.502	0.548	-9.2	111	0.00
11 T	Tert butyl alcohol	0.026	0.024	7.7	120	0.01
12 CM	1,1-Dichloroethene	0.473	0.490	-3.6#	116	0.00
13 T	Acrolein	0.023	0.021	8.7	128	-0.01
14 T	Allyl chloride	0.571	0.535	6.3	110	0.00
15 T	Acrylonitrile	0.098	0.097	1.0	111	0.00
16 T	Acetone	0.114	0.111	2.6	143	0.00
17 T	Carbon Disulfide	1.168	1.160	0.7	104	0.00
18 T	Methyl Acetate	0.365	0.397	-8.8	115	0.00
19 T	Methyl tert-butyl Ether	1.046	1.079	-3.2	115	0.00
20 T	Methylene Chloride	0.632	0.571	9.7	113	0.00
21 T	trans-1,2-Dichloroethene	0.543	0.550	-1.3	112	0.00
22 T	Diisopropyl ether	1.276	1.285	-0.7	111	0.00
23 T	Vinyl Acetate	0.488	0.467	4.3	108	0.00
24 P	1,1-Dichloroethane	0.944	0.924	2.1	113	-0.01
25 T	2-Butanone	0.136	0.126	7.4	122	0.00
26 T	2,2-Dichloropropane	0.848	0.875	-3.2	117	0.00
27 T	cis-1,2-Dichloroethene	0.644	0.645	-0.2	112	0.00
28 T	Bromochloromethane	0.338	0.334	1.2	103	0.00
29 T	Tetrahydrofuran	0.069	0.066	4.3	113	0.00
30 C	Chloroform	1.023	1.027	-0.4#	112	0.00
31 T	Cyclohexane	0.710	0.692	2.5	109	0.00
32 T	1,1,1-Trichloroethane	0.879	0.929	-5.7	117	0.00
33 S	1,2-Dichloroethane-d4	0.445	0.390	12.4	107	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	104	0.00
35 S	Dibromofluoromethane	0.322	0.306	5.0	114	0.00
36 T	1,1-Dichloropropene	0.453	0.492	-8.6	113	0.00
37 T	Ethyl Acetate	0.151	0.151	0.0	109	0.00
38 T	Carbon Tetrachloride	0.491	0.558	-13.6	121	0.00
39 T	Methylcyclohexane	0.519	0.580	-11.8	117	0.00
40 TM	Benzene	1.305	1.415	-8.4	112	0.00
41 T	Methacrylonitrile	0.095	0.100	-5.3	115	0.00
42 TM	1,2-Dichloroethane	0.353	0.366	-3.7	108	0.00
43 T	Isopropyl Acetate	0.301	0.313	-4.0	111	0.00
44 TM	Trichloroethene	0.375	0.418	-11.5	118	0.00
45 C	1,2-Dichloropropane	0.331	0.342	-3.3#	107	0.00
46 T	Dibromomethane	0.185	0.198	-7.0	109	0.00
47 T	Bromodichloromethane	0.481	0.495	-2.9	110	0.00
48 T	Methyl methacrylate	0.170	0.170	0.0	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD010224\
 Data File : VD078115.D
 Acq On : 02 Jan 2024 09:23
 Operator : RP/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 03 00:02:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122623S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 27 03:40:16 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	120	0.00
50 S	Toluene-d8	1.132	1.126	0.5	112	0.00
51 T	4-Methyl-2-Pentanone	0.155	0.163	-5.2	112	0.00
52 CM	Toluene	0.846	0.925	-9.3#	112	0.00
53 T	t-1,3-Dichloropropene	0.436	0.478	-9.6	113	0.00
54 T	cis-1,3-Dichloropropene	0.524	0.563	-7.4	112	0.00
55 T	1,1,2-Trichloroethane	0.261	0.281	-7.7	114	0.00
56 T	Ethyl methacrylate	0.207	0.231	-11.6	115	0.00
57 T	1,3-Dichloropropane	0.428	0.451	-5.4	110	0.00
58 T	2-Chloroethyl Vinyl ether	0.090	0.107	-18.9	102	0.00
59 T	2-Hexanone	0.120	0.129	-7.5	124	0.00
60 T	Dibromochloromethane	0.332	0.367	-10.5	114	0.00
61 T	1,2-Dibromoethane	0.250	0.269	-7.6	113	0.00
62 S	4-Bromofluorobenzene	0.364	0.354	2.7	111	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	104	0.00
64 T	Tetrachloroethene	0.338	0.391	-15.7	122	0.00
65 PM	Chlorobenzene	1.062	1.152	-8.5	114	0.00
66 T	1,1,1,2-Tetrachloroethane	0.402	0.455	-13.2	119	0.00
67 C	Ethyl Benzene	1.852	2.059	-11.2#	114	0.00
68 T	m/p-Xylenes	0.721	0.815	-13.0	115	0.00
69 T	o-Xylene	0.672	0.741	-10.3	114	0.00
70 T	Styrene	0.992	1.120	-12.9	111	0.00
71 P	Bromoform	0.222	0.252	-13.5	122	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	3.444	4.030	-17.0	118	0.00
74 T	N-amyl acetate	0.572	0.592	-3.5	107	0.00
75 P	1,1,2,2-Tetrachloroethane	0.626	0.668	-6.7	111	0.00
76 T	1,2,3-Trichloropropane	0.422	0.357	15.4	86	0.00
77 T	Bromobenzene	0.838	0.969	-15.6	119	0.00
78 T	n-propylbenzene	4.145	4.684	-13.0	116	0.00
79 T	2-Chlorotoluene	2.242	2.472	-10.3	113	0.00
80 T	1,3,5-Trimethylbenzene	2.848	3.260	-14.5	114	0.00
81 T	trans-1,4-Dichloro-2-butene	0.196	0.212	-8.2	115	0.00
82 T	4-Chlorotoluene	2.404	2.589	-7.7	109	0.00
83 T	tert-Butylbenzene	2.552	3.002	-17.6	120	0.00
84 T	1,2,4-Trimethylbenzene	2.901	3.274	-12.9	114	0.00
85 T	sec-Butylbenzene	3.643	4.228	-16.1	117	0.00
86 T	p-Isopropyltoluene	3.260	3.852	-18.2	121	0.00
87 T	1,3-Dichlorobenzene	1.733	1.984	-14.5	116	0.00
88 T	1,4-Dichlorobenzene	1.700	1.922	-13.1	115	0.00
89 T	n-Butylbenzene	2.992	3.394	-13.4	116	0.00
90 T	Hexachloroethane	0.652	0.744	-14.1	117	0.00
91 T	1,2-Dichlorobenzene	1.489	1.681	-12.9	115	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.088	0.094	-6.8	110	0.00
93 T	1,2,4-Trichlorobenzene	0.972	1.149	-18.2	124	0.00
94 T	Hexachlorobutadiene	0.546	0.707	-29.5#	136	0.00
95 T	Naphthalene	1.655	1.952	-17.9	121	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD010224\
Data File : VD078115.D
Acq On : 02 Jan 2024 09:23
Operator : RP/MD
Sample : VSTDCCC050
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Jan 03 00:02:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122623S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 27 03:40:16 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.825	0.999	-21.1	125	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6