



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD010224\
 Data File : VD078114.D
 Acq On : 02 Jan 2024 08:52
 Operator : RP/MD
 Sample : BFB
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122623S.M
 Title : SW846 8260
 Last Update : Wed Dec 27 03:40:16 2023



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.1	17024	PASS
75	95	30	60	49.2	49091	PASS
95	95	100	100	100.0	99728	PASS
96	95	5	9	6.8	6809	PASS
173	174	0.00	2	1.5	1496	PASS
174	95	50	100	99.1	98843	PASS
175	174	5	9	7.9	7825	PASS
176	174	95	101	95.0	93933	PASS
177	176	5	9	7.2	6730	PASS