

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_D\DATA\VD010320\
 Data File : VD064768.D
 Acq On : 04 Jan 2020 08:47
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 8.51G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 06 08:31:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 24 13:19:06 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	77	0.00
2 T	Dichlorodifluoromethane	0.389	0.349	10.3	68	0.00
3 P	Chloromethane	0.543	0.588	-8.3	84	0.00
4 C	Vinyl Chloride	0.663	0.691	-4.2#	82	0.00
5 T	Bromomethane	0.482	0.514	-6.6	86	0.00
6 T	Chloroethane	0.444	0.490	-10.4	83	0.00
7 T	Trichlorofluoromethane	1.116	1.070	4.1	74	0.00
8 T	Diethyl Ether	0.222	0.263	-18.5	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.462	0.430	6.9	71	0.00
10 T	Methyl Iodide	0.496	0.546	-10.1	76	0.00
11 T	Tert butyl alcohol	0.025	0.033	-32.0#	92	0.00
12 CM	1,1-Dichloroethene	0.440	0.443	-0.7#	77	0.00
13 T	Acrolein	0.035	0.040	-14.3	85	0.00
14 T	Allyl chloride	0.686	0.756	-10.2	84	0.00
15 T	Acrylonitrile	0.092	0.115	-25.0#	93	0.00
16 T	Acetone	0.091	0.089	2.2	75	0.00
17 T	Carbon Disulfide	1.422	1.374	3.4	75	0.00
18 T	Methyl Acetate	0.248	0.300	-21.0#	94	0.00
19 T	Methyl tert-butyl Ether	0.977	1.206	-23.4#	92	0.00
20 T	Methylene Chloride	0.524	0.569	-8.6	91	0.00
21 T	trans-1,2-Dichloroethene	0.501	0.555	-10.8	86	0.00
22 T	Diisopropyl ether	1.368	1.696	-24.0#	93	0.00
23 T	Vinyl Acetate	0.776	0.959	-23.6#	89	0.00
24 P	1,1-Dichloroethane	0.821	0.976	-18.9	90	0.00
25 T	2-Butanone	0.121	0.141	-16.5	87	0.00
26 T	2,2-Dichloropropane	0.764	0.715	6.4	72	0.00
27 T	cis-1,2-Dichloroethene	0.541	0.631	-16.6	88	0.00
28 T	Bromochloromethane	0.319	0.409	-28.2#	91	0.00
29 T	Tetrahydrofuran	0.075	0.095	-26.7#	94	0.00
30 C	Chloroform	0.862	1.052	-22.0#	94	0.00
31 T	Cyclohexane	0.822	0.742	9.7	72	0.00
32 T	1,1,1-Trichloroethane	0.809	0.901	-11.4	85	0.00
33 S	1,2-Dichloroethane-d4	0.441	0.532	-20.6#	91	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	83	0.00
35 S	Dibromofluoromethane	0.307	0.332	-8.1	86	0.00
36 T	1,1-Dichloropropene	0.481	0.475	1.2	81	0.00
37 T	Ethyl Acetate	0.189	0.216	-14.3	90	0.00
38 T	Carbon Tetrachloride	0.513	0.509	0.8	80	0.00
39 T	Methylcyclohexane	0.583	0.508	12.9	70	0.00
40 TM	Benzene	1.314	1.436	-9.3	90	0.00
41 T	Methacrylonitrile	0.110	0.125	-13.6	85	0.00
42 TM	1,2-Dichloroethane	0.380	0.438	-15.3	94	0.00
43 T	Isopropyl Acetate	0.376	0.425	-13.0	90	0.00
44 TM	Trichloroethene	0.386	0.392	-1.6	84	0.00
45 C	1,2-Dichloropropane	0.314	0.356	-13.4#	93	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_D\DATA\VD010320\
 Data File : VD064768.D
 Acq On : 04 Jan 2020 08:47
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 8.51G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 06 08:31:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 24 13:19:06 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.175	0.202	-15.4	92	0.00
47 T	Bromodichloromethane	0.465	0.530	-14.0	93	0.00
48 T	Methyl methacrylate	0.177	0.195	-10.2	84	0.00
49 T	1,4-Dioxane	0.002	0.002	0.0	96	0.00
50 S	Toluene-d8	1.175	1.252	-6.6	85	0.00
51 T	4-Methyl-2-Pentanone	0.186	0.216	-16.1	93	0.00
52 CM	Toluene	0.862	0.934	-8.4#	88	0.00
53 T	t-1,3-Dichloropropene	0.434	0.490	-12.9	91	0.00
54 T	cis-1,3-Dichloropropene	0.512	0.567	-10.7	89	0.00
55 T	1,1,2-Trichloroethane	0.236	0.267	-13.1	93	0.00
56 T	Ethyl methacrylate	0.294	0.343	-16.7	92	0.00
57 T	1,3-Dichloropropane	0.406	0.467	-15.0	93	0.00
58 T	2-Chloroethyl Vinyl ether	0.112	0.121	-8.0	93	0.00
59 T	2-Hexanone	0.126	0.147	-16.7	90	0.00
60 T	Dibromochloromethane	0.326	0.378	-16.0	94	0.00
61 T	1,2-Dibromoethane	0.232	0.254	-9.5	88	0.00
62 S	4-Bromofluorobenzene	0.372	0.410	-10.2	87	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	87	0.00
64 T	Tetrachloroethene	0.384	0.396	-3.1	89	0.00
65 PM	Chlorobenzene	1.038	1.097	-5.7	89	0.00
66 T	1,1,1,2-Tetrachloroethane	0.391	0.428	-9.5	93	0.00
67 C	Ethyl Benzene	1.884	1.957	-3.9#	88	0.00
68 T	m/p-Xylenes	0.727	0.751	-3.3	86	0.00
69 T	o-Xylene	0.657	0.697	-6.1	89	0.00
70 T	Styrene	1.153	1.279	-10.9	93	0.00
71 P	Bromoform	0.223	0.245	-9.9	92	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
73 T	Isopropylbenzene	3.573	3.645	-2.0	85	0.00
74 T	N-amyl acetate	0.764	0.859	-12.4	92	0.00
75 P	1,1,2,2-Tetrachloroethane	0.561	0.623	-11.1	94	0.00
76 T	1,2,3-Trichloropropane	0.395	0.405	-2.5	101	0.00
77 T	Bromobenzene	0.878	0.922	-5.0	89	0.00
78 T	n-propylbenzene	4.168	4.233	-1.6	85	0.00
79 T	2-Chlorotoluene	2.290	2.390	-4.4	89	0.00
80 T	1,3,5-Trimethylbenzene	2.935	3.039	-3.5	86	0.00
81 T	trans-1,4-Dichloro-2-butene	0.189	0.190	-0.5	85	0.00
82 T	4-Chlorotoluene	2.437	2.529	-3.8	88	0.00
83 T	tert-Butylbenzene	2.548	2.582	-1.3	85	0.00
84 T	1,2,4-Trimethylbenzene	2.923	3.078	-5.3	88	0.00
85 T	sec-Butylbenzene	3.494	3.477	0.5	83	0.00
86 T	p-Isopropyltoluene	3.341	3.347	-0.2	83	0.00
87 T	1,3-Dichlorobenzene	1.674	1.755	-4.8	90	0.00
88 T	1,4-Dichlorobenzene	1.644	1.716	-4.4	88	0.00
89 T	n-Butylbenzene	3.008	2.970	1.3	82	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_D\DATA\VD010320\
Data File : VD064768.D
Acq On : 04 Jan 2020 08:47
Operator : VA/SY
Sample : VSTDCCC050
Misc : 8.51G/5.00ml/MSVOA_D/SOIL
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Jan 06 08:31:57 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 24 13:19:06 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.638	0.670	-5.0	89	0.00
91 T	1,2-Dichlorobenzene	1.427	1.525	-6.9	91	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.091	0.099	-8.8	94	0.00
93 T	1,2,4-Trichlorobenzene	1.051	1.080	-2.8	86	0.00
94 T	Hexachlorobutadiene	0.661	0.597	9.7	77	0.00
95 T	Naphthalene	1.662	1.810	-8.9	89	0.00
96 T	1,2,3-Trichlorobenzene	0.897	0.952	-6.1	89	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6