

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD010322\
 Data File : VD071569.D
 Acq On : 03 Jan 2022 16:45
 Operator : VA/SY
 Sample : M5241-04
 Misc : 4.63G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-2D

Quant Time: Jan 03 16:58:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 23 08:20:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	442057	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	741494	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	699401	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	298904	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	231954	39.509	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	79.020%
35) Dibromofluoromethane	7.914	113	224864	45.122	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.240%
50) Toluene-d8	10.338	98	928527	49.305	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.600%
62) 4-Bromofluorobenzene	12.626	95	318605	47.609	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	95.220%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD010322\
Data File : VD071569.D
Acq On : 03 Jan 2022 16:45
Operator : VA/SY
Sample : M5241-04
Misc : 4.63G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-2D

Quant Time: Jan 03 16:58:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 23 08:20:24 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.979 min Scan# 1030
Delta R.T. 0.006 min
Lab File: VD071569.D
Acq: 03 Jan 2022 16:45




Instrument :
MSVOA_D
ClientSampleId :
TP-2D


Tgt Ion:168 Resp: 442057
Ion Ratio Lower Upper
168 100
99 61.0 46.6 69.8



#33
1,2-Dichloroethane-d4
Concen: 39.509 ug/l
RT: 8.332 min Scan# 1090
Delta R.T. 0.006 min
Lab File: VD071569.D
Acq: 03 Jan 2022 16:45

Tgt Ion: 65 Resp: 231954
Ion Ratio Lower Upper
65 100
67 50.6 0.0 112.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.861 min Scan# 1180

Delta R.T. 0.000 min

Lab File: VD071569.D

Acq: 03 Jan 2022 16:45

Instrument :

MSVOA_D

ClientSampleId :

TP-2D





Tgt Ion:114 Resp: 741494

Ion Ratio Lower Upper

114 100

63 23.4 0.0 41.8

88 15.1 0.0 32.8



#35

Dibromofluoromethane

Concen: 45.122 ug/l

RT: 7.914 min Scan# 1019

Delta R.T. 0.006 min

Lab File: VD071569.D

Acq: 03 Jan 2022 16:45




Tgt Ion:113 Resp: 224864


Ion Ratio Lower Upper

113 100

111 103.4 82.0 123.0

192 18.1 15.2 22.8









#50
Toluene-d8
Concen: 49.305 ug/l
RT: 10.338 min Scan# 1430
Delta R.T. 0.006 min
Lab File: VD071569.D
Acq: 03 Jan 2022 16:45




Instrument :
MSVOA_D
ClientSampleId :
TP-2D


Tgt Ion: 98 Resp: 928527
Ion Ratio Lower Upper
98 100
100 65.7 51.8 77.6



#62
4-Bromofluorobenzene
Concen: 47.609 ug/l
RT: 12.626 min Scan# 1820
Delta R.T. -0.000 min
Lab File: VD071569.D
Acq: 03 Jan 2022 16:45

Tgt Ion: 95 Resp: 318605
Ion Ratio Lower Upper
95 100
174 76.3 0.0 159.4
176 74.1 0.0 154.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.638 min Scan# 1
Delta R.T. -0.000 min
Lab File: VD071569.D
Acq: 03 Jan 2022 16:45

Instrument :
MSVOA_D
ClientSampleId :
TP-2D

Tgt Ion:117 Resp: 699401
Ion Ratio Lower Upper
117 100
82 57.1 44.7 67.1
119 31.0 26.2 39.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.567 min Scan# 1980
Delta R.T. 0.006 min
Lab File: VD071569.D
Acq: 03 Jan 2022 16:45

Tgt Ion:152 Resp: 298904
Ion Ratio Lower Upper
152 100
115 61.3 29.8 89.3
150 157.0 0.0 349.2

