

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD010421\
 Data File : VD067996.D
 Acq On : 04 Jan 2021 17:48
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 05 00:10:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122220S.M
 Quant Title : Sw846 8260
 QLast Update : Wed Dec 23 07:34:16 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	83	0.00
2 T	Dichlorodifluoromethane	0.509	0.522	-2.6	87	0.00
3 P	Chloromethane	0.696	0.619	11.1	81	0.00
4 C	Vinyl Chloride	0.649	0.591	8.9#	81	0.00
5 T	Bromomethane	0.467	0.393	15.8	80	0.00
6 T	Chloroethane	0.399	0.370	7.3	84	0.00
7 T	Trichlorofluoromethane	0.956	0.956	0.0	90	0.00
8 T	Diethyl Ether	0.274	0.290	-5.8	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.570	0.565	0.9	89	0.00
10 T	Methyl Iodide	0.498	0.517	-3.8	77	0.00
11 T	Tert butyl alcohol	0.046	0.037	19.6	101	-0.01
12 CM	1,1-Dichloroethene	0.547	0.524	4.2#	84	0.00
13 T	Acrolein	0.038	0.035	7.9	78	0.00
14 T	Allyl chloride	0.972	0.904	7.0	79	0.00
15 T	Acrylonitrile	0.131	0.132	-0.8	89	0.00
16 T	Acetone	0.103	0.101	1.9	93	0.00
17 T	Carbon Disulfide	1.905	1.768	7.2	80	0.00
18 T	Methyl Acetate	0.305	0.298	2.3	87	0.00
19 T	Methyl tert-butyl Ether	1.137	1.243	-9.3	93	0.00
20 T	Methylene Chloride	0.849	0.670	21.1	83	0.00
21 T	trans-1,2-Dichloroethene	0.653	0.650	0.5	85	0.00
22 T	Diisopropyl ether	1.958	1.977	-1.0	85	0.00
23 T	Vinyl Acetate	0.995	1.067	-7.2	88	0.00
24 P	1,1-Dichloroethane	1.185	1.157	2.4	86	0.00
25 T	2-Butanone	0.152	0.157	-3.3	91	0.00
26 T	2,2-Dichloropropane	0.934	0.928	0.6	89	0.00
27 T	cis-1,2-Dichloroethene	0.698	0.701	-0.4	86	0.00
28 T	Bromochloromethane	0.447	0.456	-2.0	85	0.00
29 T	Tetrahydrofuran	0.103	0.108	-4.9	90	0.00
30 C	Chloroform	1.161	1.191	-2.6#	88	0.00
31 T	Cyclohexane	1.104	1.004	9.1	84	0.00
32 T	1,1,1-Trichloroethane	0.994	1.000	-0.6	90	0.00
33 S	1,2-Dichloroethane-d4	0.540	0.538	0.4	92	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	81	0.00
35 S	Dibromofluoromethane	0.315	0.323	-2.5	88	0.00
36 T	1,1-Dichloropropene	0.531	0.542	-2.1	86	0.00
37 T	Ethyl Acetate	0.215	0.232	-7.9	91	0.00
38 T	Carbon Tetrachloride	0.508	0.523	-3.0	91	0.00
39 T	Methylcyclohexane	0.609	0.612	-0.5	88	0.00
40 TM	Benzene	1.542	1.569	-1.8	85	0.00
41 T	Methacrylonitrile	0.121	0.129	-6.6	89	0.00
42 TM	1,2-Dichloroethane	0.416	0.440	-5.8	92	0.00
43 T	Isopropyl Acetate	0.395	0.420	-6.3	89	0.00
44 TM	Trichloroethene	0.386	0.386	0.0	86	0.00
45 C	1,2-Dichloropropane	0.392	0.397	-1.3#	87	0.00
46 T	Dibromomethane	0.191	0.205	-7.3	93	0.00
47 T	Bromodichloromethane	0.515	0.536	-4.1	88	0.00
48 T	Methyl methacrylate	0.204	0.200	2.0	81	0.00
49 T	1,4-Dioxane	0.002	0.002	0.0	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD010421\
 Data File : VD067996.D
 Acq On : 04 Jan 2021 17:48
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 05 00:10:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122220S.M
 Quant Title : Sw846 8260
 QLast Update : Wed Dec 23 07:34:16 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
50 S	Toluene-d8	1.184	1.160	2.0	87	0.00
51 T	4-Methyl-2-Pentanone	0.204	0.228	-11.8	92	0.00
52 CM	Toluene	0.952	0.979	-2.8#	86	0.00
53 T	t-1,3-Dichloropropene	0.466	0.505	-8.4	91	0.00
54 T	cis-1,3-Dichloropropene	0.576	0.589	-2.3	86	0.00
55 T	1,1,2-Trichloroethane	0.267	0.282	-5.6	91	0.00
56 T	Ethyl methacrylate	0.300	0.344	-14.7	93	0.00
57 T	1,3-Dichloropropane	0.469	0.496	-5.8	89	0.00
58 T	2-Chloroethyl Vinyl ether	0.151	0.181	-19.9	91	0.00
59 T	2-Hexanone	0.134	0.153	-14.2	94	0.00
60 T	Dibromochloromethane	0.330	0.356	-7.9	93	0.00
61 T	1,2-Dibromoethane	0.256	0.273	-6.6	92	0.00
62 S	4-Bromofluorobenzene	0.405	0.400	1.2	87	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	83	0.00
64 T	Tetrachloroethene	0.351	0.339	3.4	89	0.00
65 PM	Chlorobenzene	1.079	1.073	0.6	88	0.00
66 T	1,1,1,2-Tetrachloroethane	0.385	0.385	0.0	88	0.00
67 C	Ethyl Benzene	1.924	1.937	-0.7#	87	0.00
68 T	m/p-Xylenes	0.733	0.755	-3.0	89	0.00
69 T	o-Xylene	0.656	0.671	-2.3	87	0.00
70 T	Styrene	1.134	1.184	-4.4	87	0.00
71 P	Bromoform	0.195	0.205	-5.1	92	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	81	0.00
73 T	Isopropylbenzene	3.711	3.852	-3.8	89	0.00
74 T	N-aryl acetate	0.842	0.886	-5.2	90	0.00
75 P	1,1,2,2-Tetrachloroethane	0.689	0.719	-4.4	94	0.00
76 T	1,2,3-Trichloropropane	0.421	0.456	-8.3	91	0.00
77 T	Bromobenzene	0.849	0.889	-4.7	89	0.00
78 T	n-propylbenzene	4.613	4.728	-2.5	88	0.00
79 T	2-Chlorotoluene	2.606	2.632	-1.0	87	0.00
80 T	1,3,5-Trimethylbenzene	3.082	3.243	-5.2	90	0.00
81 T	trans-1,4-Dichloro-2-butene	0.205	0.224	-9.3	96	0.00
82 T	4-Chlorotoluene	2.739	2.788	-1.8	88	0.00
83 T	tert-Butylbenzene	2.535	2.660	-4.9	89	0.00
84 T	1,2,4-Trimethylbenzene	3.091	3.285	-6.3	89	0.00
85 T	sec-Butylbenzene	3.709	3.850	-3.8	90	0.00
86 T	p-Isopropyltoluene	3.264	3.511	-7.6	92	0.00
87 T	1,3-Dichlorobenzene	1.688	1.718	-1.8	90	0.00
88 T	1,4-Dichlorobenzene	1.667	1.696	-1.7	91	0.00
89 T	n-Butylbenzene	3.238	3.373	-4.2	89	0.00
90 T	Hexachloroethane	0.667	0.687	-3.0	91	0.00
91 T	1,2-Dichlorobenzene	1.421	1.487	-4.6	90	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.108	0.106	1.9	96	0.00
93 T	1,2,4-Trichlorobenzene	0.844	0.921	-9.1	97	0.00
94 T	Hexachlorobutadiene	0.494	0.500	-1.2	93	0.00
95 T	Naphthalene	1.491	1.723	-15.6	98	0.00
96 T	1,2,3-Trichlorobenzene	0.729	0.804	-10.3	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD010421\
Data File : VD067996.D
Acq On : 04 Jan 2021 17:48
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Jan 05 00:10:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122220S.M
Quant Title : Sw846 8260
QLast Update : Wed Dec 23 07:34:16 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 6