

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD010422\
 Data File : VD071587.D
 Acq On : 04 Jan 2022 12:30
 Operator : VA/SY
 Sample : M5255-01
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TR-04-122921

Quant Time: Jan 04 15:41:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 23 08:20:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	229972	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	395031	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	375656	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	163844	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	176965	57.941	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.880%
35) Dibromofluoromethane	7.914	113	135618	51.082	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.160%
50) Toluene-d8	10.338	98	505555	50.389	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.780%
62) 4-Bromofluorobenzene	12.626	95	168655	47.306	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	94.620%
Target Compounds						
11) Tert butyl alcohol	5.197	59	29237	100.002	ug/l #	71
16) Acetone	4.162	43	18380	23.959	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD010422\
Data File : VD071587.D
Acq On : 04 Jan 2022 12:30
Operator : VA/SY
Sample : M5255-01
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TR-04-122921

Quant Time: Jan 04 15:41:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 23 08:20:24 2021
Response via : Initial Calibration

