

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD010422\
 Data File : VD071594.D
 Acq On : 04 Jan 2022 15:44
 Operator : VA/SY
 Sample : M5329-01
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OR-3-123021

Quant Time: Jan 05 01:28:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 23 08:20:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	6749	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	12827	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	11550	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	4257	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	7197	80.294	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	160.580%
35) Dibromofluoromethane	7.914	113	4125	47.850	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.700%
50) Toluene-d8	10.332	98	14928	45.822	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	91.640%
62) 4-Bromofluorobenzene	12.620	95	4407	38.068	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	76.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD010422\
Data File : VD071594.D
Acq On : 04 Jan 2022 15:44
Operator : VA/SY
Sample : M5329-01
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OR-3-123021

Quant Time: Jan 05 01:28:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 23 08:20:24 2021
Response via : Initial Calibration

