

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD010422\
 Data File : VD071604.D
 Acq On : 04 Jan 2022 20:21
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 05 01:31:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 23 08:20:24 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	76	0.00
2 T	Dichlorodifluoromethane	0.524	0.450	14.1	70	0.00
3 P	Chloromethane	0.671	0.623	7.2	75	0.00
4 C	Vinyl Chloride	0.804	0.714	11.2#	72	0.00
5 T	Bromomethane	0.562	0.498	11.4	76	0.02
6 T	Chloroethane	0.534	0.512	4.1	78	0.01
7 T	Trichlorofluoromethane	0.984	0.954	3.0	78	0.01
8 T	Diethyl Ether	0.282	0.298	-5.7	84	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.573	0.547	4.5	78	0.00
10 T	Methyl Iodide	0.557	0.617	-10.8	78	0.01
11 T	Tert butyl alcohol	0.065	0.034	47.7#	56	0.01
12 CM	1,1-Dichloroethene	0.521	0.491	5.8#	76	0.01
13 T	Acrolein	0.032	0.017	46.9#	41#	0.00
14 T	Allyl chloride	0.985	1.037	-5.3	83	0.02
15 T	Acrylonitrile	0.139	0.157	-12.9	86	0.00
16 T	Acetone	0.167	0.192	-15.0	83	0.00
17 T	Carbon Disulfide	1.697	1.360	19.9	65	0.00
18 T	Methyl Acetate	0.515	0.591	-14.8	94	0.00
19 T	Methyl tert-butyl Ether	1.345	1.459	-8.5	83	0.00
20 T	Methylene Chloride	0.811	0.766	5.5	93	0.00
21 T	trans-1,2-Dichloroethene	0.603	0.579	4.0	77	0.00
22 T	Diisopropyl ether	2.006	2.293	-14.3	89	0.00
23 T	Vinyl Acetate	1.034	1.158	-12.0	82	0.00
24 P	1,1-Dichloroethane	1.150	1.224	-6.4	85	0.00
25 T	2-Butanone	0.199	0.224	-12.6	84	0.00
26 T	2,2-Dichloropropane	0.945	0.984	-4.1	81	0.00
27 T	cis-1,2-Dichloroethene	0.667	0.717	-7.5	84	0.00
28 T	Bromochloromethane	0.471	0.508	-7.9	81	0.01
29 T	Tetrahydrofuran	0.120	0.133	-10.8	84	0.01
30 C	Chloroform	1.160	1.273	-9.7#	87	0.01
31 T	Cyclohexane	1.128	0.998	11.5	75	0.00
32 T	1,1,1-Trichloroethane	1.029	1.095	-6.4	84	0.00
33 S	1,2-Dichloroethane-d4	0.664	0.714	-7.5	81	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	76	0.00
35 S	Dibromofluoromethane	0.336	0.359	-6.8	81	0.00
36 T	1,1-Dichloropropene	0.537	0.520	3.2	79	0.00
37 T	Ethyl Acetate	0.248	0.278	-12.1	85	0.00
38 T	Carbon Tetrachloride	0.530	0.539	-1.7	81	0.00
39 T	Methylcyclohexane	0.628	0.568	9.6	74	0.00
40 TM	Benzene	1.443	1.481	-2.6	83	0.00
41 T	Methacrylonitrile	0.147	0.174	-18.4	82	0.00
42 TM	1,2-Dichloroethane	0.471	0.519	-10.2	89	0.00
43 T	Isopropyl Acetate	0.492	0.528	-7.3	85	0.00
44 TM	Trichloroethene	0.376	0.375	0.3	82	0.00
45 C	1,2-Dichloropropane	0.375	0.403	-7.5#	88	0.00
46 T	Dibromomethane	0.207	0.216	-4.3	82	0.00
47 T	Bromodichloromethane	0.525	0.577	-9.9	88	0.00
48 T	Methyl methacrylate	0.243	0.275	-13.2	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD010422\
 Data File : VD071604.D
 Acq On : 04 Jan 2022 20:21
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleID :
 VSTDCCC050

Quant Time: Jan 05 01:31:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 23 08:20:24 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	72	0.00
50 S	Toluene-d8	1.270	1.282	-0.9	76	0.00
51 T	4-Methyl-2-Pentanone	0.259	0.284	-9.7	85	0.00
52 CM	Toluene	0.917	0.946	-3.2#	84	0.00
53 T	t-1,3-Dichloropropene	0.491	0.530	-7.9	85	0.00
54 T	cis-1,3-Dichloropropene	0.567	0.610	-7.6	86	0.00
55 T	1,1,2-Trichloroethane	0.263	0.289	-9.9	88	0.00
56 T	Ethyl methacrylate	0.346	0.385	-11.3	84	0.00
57 T	1,3-Dichloropropane	0.481	0.523	-8.7	86	0.00
58 T	2-Chloroethyl Vinyl ether	0.166	0.187	-12.7	88	0.00
59 T	2-Hexanone	0.183	0.207	-13.1	85	0.00
60 T	Dibromochloromethane	0.330	0.364	-10.3	86	0.00
61 T	1,2-Dibromoethane	0.260	0.284	-9.2	86	0.00
62 S	4-Bromofluorobenzene	0.451	0.462	-2.4	79	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	79	0.00
64 T	Tetrachloroethene	0.347	0.327	5.8	81	0.00
65 PM	Chlorobenzene	1.026	1.032	-0.6	83	0.00
66 T	1,1,1,2-Tetrachloroethane	0.375	0.404	-7.7	87	0.00
67 C	Ethyl Benzene	1.909	1.981	-3.8#	85	0.00
68 T	m/p-Xylenes	0.725	0.751	-3.6	84	0.00
69 T	o-Xylene	0.663	0.708	-6.8	86	0.00
70 T	Styrene	1.152	1.228	-6.6	86	0.00
71 P	Bromoform	0.203	0.222	-9.4	85	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	81	0.00
73 T	Isopropylbenzene	3.855	3.823	0.8	85	0.00
74 T	N-amyl acetate	1.034	1.070	-3.5	83	0.00
75 P	1,1,2,2-Tetrachloroethane	0.697	0.734	-5.3	88	0.00
76 T	1,2,3-Trichloropropane	0.594	0.528	11.1	76	0.00
77 T	Bromobenzene	0.858	0.836	2.6	84	0.00
78 T	n-propylbenzene	4.832	4.730	2.1	83	0.00
79 T	2-Chlorotoluene	2.700	2.692	0.3	85	0.00
80 T	1,3,5-Trimethylbenzene	3.180	3.255	-2.4	86	0.00
81 T	trans-1,4-Dichloro-2-butene	0.210	0.199	5.2	80	0.00
82 T	4-Chlorotoluene	2.801	2.830	-1.0	86	0.00
83 T	tert-Butylbenzene	2.717	2.693	0.9	83	0.00
84 T	1,2,4-Trimethylbenzene	3.147	3.167	-0.6	85	0.00
85 T	sec-Butylbenzene	4.142	4.162	-0.5	85	0.00
86 T	p-Isopropyltoluene	3.437	3.454	-0.5	85	0.00
87 T	1,3-Dichlorobenzene	1.727	1.707	1.2	86	0.00
88 T	1,4-Dichlorobenzene	1.694	1.656	2.2	85	0.00
89 T	n-Butylbenzene	3.373	3.353	0.6	84	0.00
90 T	Hexachloroethane	0.651	0.655	-0.6	85	0.00
91 T	1,2-Dichlorobenzene	1.466	1.486	-1.4	86	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.121	0.122	-0.8	90	0.00
93 T	1,2,4-Trichlorobenzene	0.893	0.887	0.7	83	0.00
94 T	Hexachlorobutadiene	0.479	0.473	1.3	84	0.00
95 T	Naphthalene	1.701	1.740	-2.3	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD010422\
Data File : VD071604.D
Acq On : 04 Jan 2022 20:21
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Jan 05 01:31:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 23 08:20:24 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.773	0.752	2.7	82	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6