

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD010522\
 Data File : VD071621.D
 Acq On : 05 Jan 2022 18:29
 Operator : VA/SY
 Sample : M5236-06RE
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-3-(19.5-20)RE

Quant Time: Jan 06 07:05:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 23 08:20:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	88806	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	156536	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	149151	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	66361	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	84637	71.761	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	143.520%
35) Dibromofluoromethane	7.914	113	56467	53.674	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.340%
50) Toluene-d8	10.338	98	196299	49.375	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.740%
62) 4-Bromofluorobenzene	12.626	95	65284	46.210	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD010522\
Data File : VD071621.D
Acq On : 05 Jan 2022 18:29
Operator : VA/SY
Sample : M5236-06RE
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-3-(19.5-20)RE

Quant Time: Jan 06 07:05:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122221S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 23 08:20:24 2021
Response via : Initial Calibration

