

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_D\DATA\VD010720\
 Data File : VD064796.D
 Acq On : 07 Jan 2020 16:45
 Operator : VA/SY
 Sample : L1049-07
 Misc : 6.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-3

Quant Time: Jan 08 08:17:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 24 13:19:06 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	481148	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	750996	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	632133	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	261412	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	195855	46.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.32%	
35) Dibromofluoromethane	7.92	113	220980	47.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.92%	
50) Toluene-d8	10.34	98	882335	50.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.00%	
62) 4-Bromofluorobenzene	12.64	95	240413	43.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.18%	
Target Compounds						
16) Acetone	4.17	43	14712	18.735	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_D\DATA\VD010720\
Data File : VD064796.D
Acq On : 07 Jan 2020 16:45
Operator : VA/SY
Sample : L1049-07
Misc : 6.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-3

Quant Time: Jan 08 08:17:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 24 13:19:06 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.98 min Scan# 1030

Delta R.T. -0.01 min

Lab File: VD064796.D

Acq: 07 Jan 2020 16:45

Instrument :

MSVOA_D

ClientSampleId :


TP-3

Tgt Ion:168 Resp: 481148

Ion Ratio Lower Upper

168 100

99 51.1 41.8 62.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VDC

7.98

250000

200000



150000

100000

50000

0

Time-->



#16

Acetone

Concen: 18.735 ug/l

RT: 4.17 min Scan# 382

Delta R.T. 0.00 min

Lab File: VD064796.D


Acq: 07 Jan 2020 16:45

Tgt Ion: 43 Resp: 14712

Ion Ratio Lower Upper

43 100

58 27.7 25.8 38.6



Abundance Ion 43.00 (42.70 to 43.70): VDC

Ion 58.00 (57.70 to 58.70): VDC

4.17

6000

5000

4000


3000


2000

1000

0

Time-->





#33

1,2-Dichloroethane-d4

Concen: 46.163 ug/l

RT: 8.33 min Scan# 1090

Delta R.T. 0.00 min

Lab File: VD064796.D

Acq: 07 Jan 2020 16:45

Instrument :

MSVOA_D

ClientSampleId :


TP-3

Tgt Ion: 65 Resp: 195855

Ion Ratio Lower Upper




65 100

67 50.2 0.0 106.2



Abundance Ion 64.90 (64.60 to 65.60): VDC

Ion 66.90 (66.60 to 67.60): VDC



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.87 min Scan# 1181

Delta R.T. 0.00 min

Lab File: VD064796.D

Acq: 07 Jan 2020 16:45


Tgt Ion: 114 Resp: 750996

Ion Ratio Lower Upper

114 100

63 19.8 0.0 38.0



88 14.2 0.0 29.2



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VDC

Ion 87.90 (87.60 to 88.60): VDC









#35
 Dibromofluoromethane
 Concen: 47.961 ug/l
 RT: 7.92 min Scan# 1020
 Delta R.T. 0.00 min
 Lab File: VD064796.D
 Acq: 07 Jan 2020 16:45




Instrument :
 MSVOA_D
 ClientSampleId :
 TP-3


Tgt Ion: 113 Resp: 220980
 Ion Ratio Lower Upper
 113 100
 111 102.3 80.6 121.0
 192 21.7 17.9 26.9



#50
 Toluene-d8
 Concen: 50.001 ug/l
 RT: 10.34 min Scan# 1432
 Delta R.T. 0.01 min
 Lab File: VD064796.D
 Acq: 07 Jan 2020 16:45

Tgt Ion: 98 Resp: 882335
 Ion Ratio Lower Upper
 98 100
 100 64.9 51.7 77.5









#62
4-Bromofluorobenzene
Concen: 43.085 ug/l
RT: 12.64 min Scan# 1822
Delta R.T. 0.00 min
Lab File: VD064796.D
Acq: 07 Jan 2020 16:45




Instrument :
MSVOA_D
ClientSampleId :
TP-3


Tgt Ion: 95 Resp: 240413
Ion Ratio Lower Upper
95 100
174 94.9 0.0 193.2
176 92.5 0.0 184.2



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.65 min Scan# 1654
Delta R.T. 0.00 min
Lab File: VD064796.D
Acq: 07 Jan 2020 16:45

Tgt Ion: 117 Resp: 632133
Ion Ratio Lower Upper
117 100
82 54.6 42.2 63.4
119 33.5 26.0 39.0





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.58 min Scan# 1983

Delta R.T. 0.00 min

Lab File: VD064796.D

Acq: 07 Jan 2020 16:45

Instrument :

MSVOA_D

ClientSampleId :

TP-3


Tgt Ion:152 Resp: 261412

Ion Ratio Lower Upper

152 100

115 54.7 27.8 83.4

150 154.5 0.0 351.4



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

