


Dichlorodifluoromethane

Response Ratio



$$\text{Response} = 3.748\text{e-}001 * \text{Amt} + 4.139\text{e-}003$$


Coef of Det (r^2) = 0.996495 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D010722S.M

Calibration Table Last Updated: Fri Jan 07 15:39:14 2022

Chloromethane

Response Ratio



$$\text{Response} = 4.864\text{e-}001 * \text{Amt} + 1.035\text{e-}002$$


Coef of Det (r^2) = 0.998696 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D010722S.M

Calibration Table Last Updated: Fri Jan 07 15:39:14 2022

Bromomethane

Response Ratio



$$\text{Response} = 4.095\text{e-}001 * \text{Amt} + 5.777\text{e-}003$$


Coef of Det (r^2) = 0.993522 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D010722S.M

Calibration Table Last Updated: Fri Jan 07 15:39:14 2022

Tert butyl alcohol


Response Ratio



$R = 7.362e-004 A^2 + 1.549e-002 A + 6.533e-002$
 Coef of Det (r^2) = 0.999304 Curve Fit: Quadratic
 Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D010722S.M
 Calibration Table Last Updated: Fri Jan 07 15:39:14 2022

Carbon Disulfide

Response Ratio



$$\text{Response} = 1.133\text{e}+000 * \text{Amt} - 1.075\text{e}-002$$


Coef of Det (r^2) = 0.997003 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D010722S.M

Calibration Table Last Updated: Fri Jan 07 15:39:14 2022

Methyl Acetate

Response Ratio



$$\text{Response} = 4.229\text{e-}001 * \text{Amt} + 3.599\text{e-}003$$


Coef of Det (r^2) = 0.992687 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D010722S.M

Calibration Table Last Updated: Fri Jan 07 15:39:14 2022

Methylene Chloride

Response Ratio



$$\text{Response} = 5.204\text{e-}001 * \text{Amt} + 2.885\text{e-}002$$


Coef of Det (r^2) = 0.996291 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D010722S.M

Calibration Table Last Updated: Fri Jan 07 15:39:14 2022

Cyclohexane

Response Ratio




$$\text{Response} = 8.963\text{e-}001 * \text{Amt} - 5.753\text{e-}003$$

Coef of Det (r^2) = 0.995453 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D010722S.M

Calibration Table Last Updated: Fri Jan 07 15:39:14 2022

Methacrylonitrile



Response = 1.380e-001 * Amt - 3.120e-003
Coef of Det (r^2) = 0.991664 Curve Fit: Linear
Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D010722S.M
Calibration Table Last Updated: Fri Jan 07 15:39:14 2022