



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD010722\
 Data File : VD071639.D
 Acq On : 07 Jan 2022 09:23
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
 Title : SW846 8260
 Last Update : Fri Jan 07 15:39:14 2022



AutoFind: Scans 1819, 1820, 1821; Background Corrected with Scan 1811

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	25.7	61579	PASS
75	95	30	60	58.6	140373	PASS
95	95	100	100	100.0	239467	PASS
96	95	5	9	6.6	15881	PASS
173	174	0.00	2	2.0	3620	PASS
174	95	50	100	75.9	181760	PASS
175	174	5	9	8.0	14533	PASS
176	174	95	101	98.0	178091	PASS
177	176	5	9	6.7	11907	PASS