

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_D\DATA\VD010820\
 Data File : VD064804.D
 Acq On : 08 Jan 2020 14:10
 Operator : VA/SY
 Sample : DI01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 DI01

Manual Integrations
 APPROVED

apatel
 1/9/2020 3:45:43 PM

Quant Time: Jan 08 14:59:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 24 13:19:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	515367	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	810334	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	724318	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	341126	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	252603	55.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.18%	
35) Dibromofluoromethane	7.92	113	252476	50.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.56%	
50) Toluene-d8	10.34	98	968566	50.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.74%	
62) 4-Bromofluorobenzene	12.64	95	299410	49.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.46%	
Target Compounds						
16) Acetone	4.17	43	6467	10.139	ug/l #	80
20) Methylene Chloride	4.96	84	17536	Below Cal		96
29) Tetrahydrofuran	7.58	42	2922m	3.761	ug/l	
52) Toluene	10.40	92	15413	1.103	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_D\DATA\VD010820\
Data File : VD064804.D
Acq On : 08 Jan 2020 14:10
Operator : VA/SY
Sample : DI01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
DI01

Manual Integrations
APPROVED

apatel
1/9/2020 3:45:43 PM

Quant Time: Jan 08 14:59:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 24 13:19:06 2019
Response via : Initial Calibration

