

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD010920\
 Data File : VD064819.D
 Acq On : 09 Jan 2020 16:17
 Operator : VA/SY
 Sample : L1051-06
 Misc : 7.99G/5.00ml/MSVOA D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-3F-(8.5-10.5)

Quant Time: Jan 10 05:49:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 24 13:19:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	463733	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	725176	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	626067	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	279589	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	191551	46.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.68%	
35) Dibromofluoromethane	7.92	113	205473	46.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.36%	
50) Toluene-d8	10.34	98	869452	51.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.06%	
62) 4-Bromofluorobenzene	12.64	95	251856	46.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.48%	
Target Compounds						
16) Acetone	4.17	43	4060	8.339	ug/l #	85
30) Chloroform	7.71	83	9553	1.195	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD010920\
Data File : VD064819.D
Acq On : 09 Jan 2020 16:17
Operator : VA/SY
Sample : L1051-06
Misc : 7.99G/5.00ml/MSVOA D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-3F-(8.5-10.5)

Quant Time: Jan 10 05:49:48 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 24 13:19:06 2019
Response via : Initial Calibration

