

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD010920\
 Data File : VD064820.D
 Acq On : 09 Jan 2020 16:59
 Operator : VA/SY
 Sample : L1051-07
 Misc : 7.91G/5.00ml/MSVOA D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-3G-(11-13)

Quant Time: Jan 10 05:42:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 24 13:19:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	0.00	168	0	0.00	ug/l	-7.98
34) 1,4-Difluorobenzene	0.00	114	0	0.00	ug/l	-8.87
63) Chlorobenzene-d5	11.65	117	155	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	71	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.57	65	55	0.00	ug/l	0.24
Spiked Amount 50.000			Recovery	=	0.00%	
35) Dibromofluoromethane	0.00	113	0	0.00	ug/l	
Spiked Amount 50.000			Recovery	=	0.00%	
50) Toluene-d8	10.34	98	118	0.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	0.00%	
62) 4-Bromofluorobenzene	12.52	95	60	0.00	ug/l	-0.12
Spiked Amount 50.000			Recovery	=	0.00%	
Target Compounds						
					Qvalue	
74) N-amyl acetate	12.31	43	90	82.980	ug/l #	21
78) n-propylbenzene	12.82	91	54	9.123	ug/l #	51
79) 2-Chlorotoluene	12.91	91	56	17.222	ug/l #	39
82) 4-Chlorotoluene	12.91	91	56	16.183	ug/l #	39
87) 1,3-Dichlorobenzene	13.64	146	66	27.770	ug/l #	19
88) 1,4-Dichlorobenzene	13.64	146	66	28.269	ug/l #	18
91) 1,2-Dichlorobenzene	13.64	146	66	32.567	ug/l #	20
93) 1,2,4-Trichlorobenzene	15.29	180	57	38.199	ug/l #	12

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD010920\
Data File : VD064820.D
Acq On : 09 Jan 2020 16:59
Operator : VA/SY
Sample : L1051-07
Misc : 7.91G/5.00ml/MSVOA D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-3G-(11-13)

Quant Time: Jan 10 05:42:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 24 13:19:06 2019
Response via : Initial Calibration

