

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD011020\
 Data File : VD064835.D
 Acq On : 10 Jan 2020 16:08
 Operator : VA/SY
 Sample : L1063-06
 Misc : 4.94G/5.00ml/MSVOA D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WP022SB452_200108_FD_30-32

Quant Time: Jan 10 23:04:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 24 13:19:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	439604	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	662562	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	471090	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	145542	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	173994	44.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.78%	
35) Dibromofluoromethane	7.92	113	196828	48.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.84%	
50) Toluene-d8	10.34	98	737567	47.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.76%	
62) 4-Bromofluorobenzene	12.64	95	164402	33.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	66.80%	
Target Compounds						
4) Vinyl Chloride	2.35	62	7926	1.361	ug/l	89
16) Acetone	4.17	43	8828	13.723	ug/l #	85
17) Carbon Disulfide	4.40	76	14304	1.144	ug/l	97
27) cis-1,2-Dichloroethene	7.21	96	6111	1.286	ug/l	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD011020\
Data File : VD064835.D
Acq On : 10 Jan 2020 16:08
Operator : VA/SY
Sample : L1063-06
Misc : 4.94G/5.00ml/MSVOA D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WP022SB452_200108_FD_30-32

Quant Time: Jan 10 23:04:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 24 13:19:06 2019
Response via : Initial Calibration

