

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD011020\
 Data File : VD064838.D
 Acq On : 10 Jan 2020 17:35
 Operator : VA/SY
 Sample : L1063-05
 Misc : 6.64G/5.00ml/MSVOA D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WP022SB452_200108_30-32

Quant Time: Jan 13 05:52:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 24 13:19:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	504613	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	760241	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	599293	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	206944	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	207250	46.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.16%	
35) Dibromofluoromethane	7.91	113	224679	48.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.34%	
50) Toluene-d8	10.34	98	886657	49.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.28%	
62) 4-Bromofluorobenzene	12.64	95	222256	39.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.70%	
Target Compounds						
4) Vinyl Chloride	2.35	62	12065	1.804	ug/l	95
16) Acetone	4.16	43	12521	15.980	ug/l	96
18) Methyl Acetate	4.71	43	7400	2.954	ug/l	98
27) cis-1,2-Dichloroethene	7.21	96	20163	3.695	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD011020\
Data File : VD064838.D
Acq On : 10 Jan 2020 17:35
Operator : VA/SY
Sample : L1063-05
Misc : 6.64G/5.00ml/MSVOA D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WP022SB452_200108_30-32

Quant Time: Jan 13 05:52:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 24 13:19:06 2019
Response via : Initial Calibration

