

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD011020\
 Data File : VD064839.D
 Acq On : 10 Jan 2020 18:30
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 10 23:12:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 24 13:19:06 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	91	0.00
2 T	Dichlorodifluoromethane	0.389	0.359	7.7	83	0.00
3 P	Chloromethane	0.543	0.563	-3.7	95	0.00
4 C	Vinyl Chloride	0.663	0.658	0.8#	92	0.00
5 T	Bromomethane	0.482	0.448	7.1	88	0.00
6 T	Chloroethane	0.444	0.457	-2.9	92	0.00
7 T	Trichlorofluoromethane	1.116	1.101	1.3	90	0.00
8 T	Diethyl Ether	0.222	0.235	-5.9	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.462	0.489	-5.8	95	0.00
10 T	Methyl Iodide	0.496	0.517	-4.2	86	0.00
11 T	Tert butyl alcohol	0.025	0.030	-20.0	100	0.00
12 CM	1,1-Dichloroethene	0.440	0.447	-1.6#	92	0.00
13 T	Acrolein	0.035	0.034	2.9	84	0.00
14 T	Allyl chloride	0.686	0.752	-9.6	99	0.00
15 T	Acrylonitrile	0.092	0.105	-14.1	101	0.00
16 T	Acetone	0.091	0.119	-30.8#	118	0.00
17 T	Carbon Disulfide	1.422	1.289	9.4	83	0.00
18 T	Methyl Acetate	0.248	0.275	-10.9	101	0.00
19 T	Methyl tert-butyl Ether	0.977	1.102	-12.8	99	0.00
20 T	Methylene Chloride	0.524	0.519	1.0	99	0.00
21 T	trans-1,2-Dichloroethene	0.501	0.518	-3.4	95	0.00
22 T	Diisopropyl ether	1.368	1.556	-13.7	101	0.00
23 T	Vinyl Acetate	0.776	0.874	-12.6	96	0.00
24 P	1,1-Dichloroethane	0.821	0.926	-12.8	102	0.00
25 T	2-Butanone	0.121	0.146	-20.7#	106	0.00
26 T	2,2-Dichloropropane	0.764	0.829	-8.5	99	0.00
27 T	cis-1,2-Dichloroethene	0.541	0.593	-9.6	98	0.00
28 T	Bromochloromethane	0.319	0.366	-14.7	96	0.00
29 T	Tetrahydrofuran	0.075	0.085	-13.3	99	0.00
30 C	Chloroform	0.862	0.953	-10.6#	101	0.00
31 T	Cyclohexane	0.822	0.805	2.1	93	0.00
32 T	1,1,1-Trichloroethane	0.809	0.882	-9.0	99	0.00
33 S	1,2-Dichloroethane-d4	0.441	0.481	-9.1	98	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	91	0.00
35 S	Dibromofluoromethane	0.307	0.332	-8.1	94	0.00
36 T	1,1-Dichloropropene	0.481	0.502	-4.4	94	0.00
37 T	Ethyl Acetate	0.189	0.212	-12.2	97	0.00
38 T	Carbon Tetrachloride	0.513	0.549	-7.0	95	0.00
39 T	Methylcyclohexane	0.583	0.588	-0.9	89	0.00
40 TM	Benzene	1.314	1.418	-7.9	98	0.00
41 T	Methacrylonitrile	0.110	0.119	-8.2	89	-0.01
42 TM	1,2-Dichloroethane	0.380	0.427	-12.4	101	0.00
43 T	Isopropyl Acetate	0.376	0.416	-10.6	97	0.00
44 TM	Trichloroethene	0.386	0.406	-5.2	95	0.00
45 C	1,2-Dichloropropane	0.314	0.351	-11.8#	101	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD011020\
 Data File : VD064839.D
 Acq On : 10 Jan 2020 18:30
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 10 23:12:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 24 13:19:06 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.175	0.193	-10.3	97	0.00
47 T	Bromodichloromethane	0.465	0.517	-11.2	99	0.00
48 T	Methyl methacrylate	0.177	0.195	-10.2	92	0.00
49 T	1,4-Dioxane	0.002	0.002	0.0	100	0.01
50 S	Toluene-d8	1.175	1.248	-6.2	93	0.00
51 T	4-Methyl-2-Pentanone	0.186	0.213	-14.5	101	0.00
52 CM	Toluene	0.862	0.931	-8.0#	97	0.00
53 T	t-1,3-Dichloropropene	0.434	0.487	-12.2	99	0.00
54 T	cis-1,3-Dichloropropene	0.512	0.572	-11.7	98	0.00
55 T	1,1,2-Trichloroethane	0.236	0.258	-9.3	98	0.00
56 T	Ethyl methacrylate	0.294	0.333	-13.3	99	0.00
57 T	1,3-Dichloropropane	0.406	0.450	-10.8	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.112	0.116	-3.6	98	0.00
59 T	2-Hexanone	0.126	0.156	-23.8#	105	0.00
60 T	Dibromochloromethane	0.326	0.357	-9.5	98	0.00
61 T	1,2-Dibromoethane	0.232	0.251	-8.2	96	0.00
62 S	4-Bromofluorobenzene	0.372	0.404	-8.6	95	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	93	0.00
64 T	Tetrachloroethene	0.384	0.434	-13.0	104	0.00
65 PM	Chlorobenzene	1.038	1.115	-7.4	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.391	0.437	-11.8	101	0.00
67 C	Ethyl Benzene	1.884	2.014	-6.9#	97	0.00
68 T	m/p-Xylenes	0.727	0.784	-7.8	96	0.00
69 T	o-Xylene	0.657	0.712	-8.4	97	0.00
70 T	Styrene	1.153	1.268	-10.0	99	0.00
71 P	Bromoform	0.223	0.243	-9.0	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
73 T	Isopropylbenzene	3.573	3.890	-8.9	96	0.00
74 T	N-amyl acetate	0.764	0.878	-14.9	100	0.00
75 P	1,1,2,2-Tetrachloroethane	0.561	0.624	-11.2	99	0.00
76 T	1,2,3-Trichloropropane	0.395	0.406	-2.8	107	0.00
77 T	Bromobenzene	0.878	0.951	-8.3	97	0.00
78 T	n-propylbenzene	4.168	4.560	-9.4	97	0.00
79 T	2-Chlorotoluene	2.290	2.538	-10.8	99	0.00
80 T	1,3,5-Trimethylbenzene	2.935	3.193	-8.8	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.189	0.209	-10.6	99	0.00
82 T	4-Chlorotoluene	2.437	2.636	-8.2	97	0.00
83 T	tert-Butylbenzene	2.548	2.759	-8.3	95	0.00
84 T	1,2,4-Trimethylbenzene	2.923	3.223	-10.3	98	0.00
85 T	sec-Butylbenzene	3.494	3.856	-10.4	98	0.00
86 T	p-Isopropyltoluene	3.341	3.696	-10.6	97	0.00
87 T	1,3-Dichlorobenzene	1.674	1.816	-8.5	98	0.00
88 T	1,4-Dichlorobenzene	1.644	1.789	-8.8	97	0.00
89 T	n-Butylbenzene	3.008	3.320	-10.4	97	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD011020\
Data File : VD064839.D
Acq On : 10 Jan 2020 18:30
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Jan 10 23:12:48 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D122419S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 24 13:19:06 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.638	0.698	-9.4	97	0.00
91 T	1,2-Dichlorobenzene	1.427	1.533	-7.4	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.091	0.103	-13.2	103	0.00
93 T	1,2,4-Trichlorobenzene	1.051	1.098	-4.5	92	0.00
94 T	Hexachlorobutadiene	0.661	0.655	0.9	89	0.00
95 T	Naphthalene	1.662	1.813	-9.1	94	0.00
96 T	1,2,3-Trichlorobenzene	0.897	0.957	-6.7	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6