

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011022\
 Data File : VD071649.D
 Acq On : 10 Jan 2022 09:01
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 11 04:05:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 07 15:39:14 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	107	0.00
2 T	Dichlorodifluoromethane	0.411	0.397	3.4	117	0.00
3 P	Chloromethane	0.541	0.516	4.6	115	0.00
4 C	Vinyl Chloride	0.622	0.606	2.6#	116	0.00
5 T	Bromomethane	0.467	0.421	9.9	120	0.00
6 T	Chloroethane	0.443	0.414	6.5	112	0.00
7 T	Trichlorofluoromethane	0.881	0.878	0.3	120	0.00
8 T	Diethyl Ether	0.232	0.231	0.4	114	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.515	0.534	-3.7	122	0.00
10 T	Methyl Iodide	0.493	0.481	2.4	104	0.00
11 T	Tert butyl alcohol	0.061	0.034	44.3#	116	0.00
12 CM	1,1-Dichloroethene	0.447	0.456	-2.0#	120	0.00
13 T	Acrolein	0.063	0.051	19.0	87	0.00
14 T	Allyl chloride	0.886	0.905	-2.1	122	0.00
15 T	Acrylonitrile	0.122	0.121	0.8	112	0.00
16 T	Acetone	0.149	0.160	-7.4	119	0.00
17 T	Carbon Disulfide	1.186	1.142	3.7	115	0.00
18 T	Methyl Acetate	0.465	0.408	12.3	108	0.00
19 T	Methyl tert-butyl Ether	1.180	1.187	-0.6	115	0.00
20 T	Methylene Chloride	0.652	0.541	17.0	113	0.00
21 T	trans-1,2-Dichloroethene	0.510	0.509	0.2	117	0.00
22 T	Diisopropyl ether	1.908	1.938	-1.6	113	0.00
23 T	Vinyl Acetate	0.891	0.958	-7.5	117	0.00
24 P	1,1-Dichloroethane	1.089	1.100	-1.0	117	0.00
25 T	2-Butanone	0.175	0.172	1.7	113	0.00
26 T	2,2-Dichloropropane	0.915	0.967	-5.7	123	0.00
27 T	cis-1,2-Dichloroethene	0.602	0.615	-2.2	115	0.00
28 T	Bromochloromethane	0.531	0.530	0.2	108	0.00
29 T	Tetrahydrofuran	0.101	0.098	3.0	111	0.00
30 C	Chloroform	1.114	1.125	-1.0#	117	0.00
31 T	Cyclohexane	0.959	0.917	4.4	118	0.00
32 T	1,1,1-Trichloroethane	0.982	1.003	-2.1	119	0.00
33 S	1,2-Dichloroethane-d4	0.643	0.645	-0.3	103	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	107	0.00
35 S	Dibromofluoromethane	0.330	0.321	2.7	103	0.00
36 T	1,1-Dichloropropene	0.471	0.481	-2.1	119	0.00
37 T	Ethyl Acetate	0.219	0.212	3.2	110	0.00
38 T	Carbon Tetrachloride	0.477	0.504	-5.7	122	0.00
39 T	Methylcyclohexane	0.509	0.536	-5.3	120	0.00
40 TM	Benzene	1.299	1.307	-0.6	115	0.00
41 T	Methacrylonitrile	0.143	0.122	14.7	119	0.00
42 TM	1,2-Dichloroethane	0.440	0.441	-0.2	116	0.00
43 T	Isopropyl Acetate	0.412	0.417	-1.2	115	0.00
44 TM	Trichloroethene	0.334	0.334	0.0	116	0.00
45 C	1,2-Dichloropropane	0.351	0.354	-0.9#	115	0.00
46 T	Dibromomethane	0.180	0.179	0.6	113	0.00
47 T	Bromodichloromethane	0.493	0.507	-2.8	116	0.00
48 T	Methyl methacrylate	0.208	0.209	-0.5	123	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011022\
 Data File : VD071649.D
 Acq On : 10 Jan 2022 09:01
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 11 04:05:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 07 15:39:14 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	100	0.00
50 S	Toluene-d8	1.237	1.277	-3.2	104	0.00
51 T	4-Methyl-2-Pentanone	0.218	0.221	-1.4	113	0.00
52 CM	Toluene	0.817	0.839	-2.7#	116	0.00
53 T	t-1,3-Dichloropropene	0.433	0.443	-2.3	118	0.00
54 T	cis-1,3-Dichloropropene	0.510	0.513	-0.6	114	0.00
55 T	1,1,2-Trichloroethane	0.236	0.241	-2.1	114	0.00
56 T	Ethyl methacrylate	0.289	0.302	-4.5	114	0.00
57 T	1,3-Dichloropropane	0.435	0.431	0.9	115	0.00
58 T	2-Chloroethyl Vinyl ether	0.159	0.165	-3.8	115	0.00
59 T	2-Hexanone	0.152	0.162	-6.6	118	0.00
60 T	Dibromochloromethane	0.300	0.310	-3.3	116	0.00
61 T	1,2-Dibromoethane	0.225	0.225	0.0	113	0.00
62 S	4-Bromofluorobenzene	0.436	0.434	0.5	106	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	105	0.00
64 T	Tetrachloroethene	0.294	0.306	-4.1	118	0.00
65 PM	Chlorobenzene	0.957	0.974	-1.8	115	0.00
66 T	1,1,1,2-Tetrachloroethane	0.363	0.364	-0.3	117	0.00
67 C	Ethyl Benzene	1.757	1.839	-4.7#	115	0.00
68 T	m/p-Xylenes	0.653	0.702	-7.5	117	0.00
69 T	o-Xylene	0.618	0.655	-6.0	118	0.00
70 T	Styrene	1.060	1.128	-6.4	115	0.00
71 P	Bromoform	0.177	0.187	-5.6	117	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
73 T	Isopropylbenzene	3.593	3.893	-8.3	117	0.00
74 T	N-amyl acetate	0.881	0.910	-3.3	112	0.00
75 P	1,1,2,2-Tetrachloroethane	0.637	0.633	0.6	112	0.00
76 T	1,2,3-Trichloropropane	0.483	0.442	8.5	113	0.00
77 T	Bromobenzene	0.776	0.802	-3.4	116	0.00
78 T	n-propylbenzene	4.519	4.909	-8.6	118	0.00
79 T	2-Chlorotoluene	2.547	2.691	-5.7	116	0.00
80 T	1,3,5-Trimethylbenzene	3.002	3.289	-9.6	118	0.00
81 T	trans-1,4-Dichloro-2-butene	0.173	0.180	-4.0	118	0.00
82 T	4-Chlorotoluene	2.642	2.824	-6.9	116	0.00
83 T	tert-Butylbenzene	2.565	2.789	-8.7	118	0.00
84 T	1,2,4-Trimethylbenzene	2.969	3.196	-7.6	116	0.00
85 T	sec-Butylbenzene	3.929	4.296	-9.3	115	0.00
86 T	p-Isopropyltoluene	3.226	3.579	-10.9	119	0.00
87 T	1,3-Dichlorobenzene	1.602	1.686	-5.2	116	0.00
88 T	1,4-Dichlorobenzene	1.634	1.654	-1.2	114	0.00
89 T	n-Butylbenzene	3.231	3.536	-9.4	118	0.00
90 T	Hexachloroethane	0.638	0.689	-8.0	121	0.00
91 T	1,2-Dichlorobenzene	1.373	1.432	-4.3	115	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.102	0.103	-1.0	118	0.00
93 T	1,2,4-Trichlorobenzene	0.791	0.842	-6.4	115	0.00
94 T	Hexachlorobutadiene	0.464	0.498	-7.3	119	0.00
95 T	Naphthalene	1.401	1.472	-5.1	111	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011022\
Data File : VD071649.D
Acq On : 10 Jan 2022 09:01
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Jan 11 04:05:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 07 15:39:14 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.674	0.706	-4.7	114	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6