

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011022\
 Data File : VD071663.D
 Acq On : 10 Jan 2022 17:55
 Operator : VA/SY
 Sample : N1092-01
 Misc : 5.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OK-03-011022

Quant Time: Jan 11 04:10:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 07 15:39:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	341075	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	624651	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	566431	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	224966	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	220418	50.224	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	100.440%
35) Dibromofluoromethane	7.914	113	197791	47.913	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.820%
50) Toluene-d8	10.338	98	807365	52.246	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.500%
62) 4-Bromofluorobenzene	12.626	95	238395	43.780	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	87.560%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011022\
Data File : VD071663.D
Acq On : 10 Jan 2022 17:55
Operator : VA/SY
Sample : N1092-01
Misc : 5.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OK-03-011022

Quant Time: Jan 11 04:10:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 07 15:39:14 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.979 min Scan# 1030
Delta R.T. 0.000 min
Lab File: VD071663.D
Acq: 10 Jan 2022 17:55




Instrument :
MSVOA_D
ClientSampleId :
OK-03-011022


Tgt Ion:168 Resp: 341075
Ion Ratio Lower Upper
168 100
99 61.8 51.8 77.6



#33
1,2-Dichloroethane-d4
Concen: 50.224 ug/l
RT: 8.326 min Scan# 1089
Delta R.T. -0.006 min
Lab File: VD071663.D
Acq: 10 Jan 2022 17:55

Tgt Ion: 65 Resp: 220418
Ion Ratio Lower Upper
65 100
67 51.1 0.0 101.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.861 min Scan# 1180

Delta R.T. -0.006 min

Lab File: VD071663.D

Acq: 10 Jan 2022 17:55

Instrument :

MSVOA_D

ClientSampleId :

OK-03-011022





Tgt Ion:114 Resp: 624651

Ion Ratio Lower Upper

114 100

63 24.2 0.0 46.2

88 16.3 0.0 32.0



#35

Dibromofluoromethane

Concen: 47.913 ug/l

RT: 7.914 min Scan# 1019

Delta R.T. 0.000 min

Lab File: VD071663.D

Acq: 10 Jan 2022 17:55




Tgt Ion:113 Resp: 197791


Ion Ratio Lower Upper

113 100

111 101.7 82.2 123.4

192 17.6 14.2 21.2









#50
Toluene-d8
Concen: 52.246 ug/l
RT: 10.338 min Scan# 1431
Delta R.T. 0.001 min
Lab File: VD071663.D
Acq: 10 Jan 2022 17:55




Instrument :
MSVOA_D
ClientSampleId :
OK-03-011022


Tgt Ion: 98 Resp: 807365
Ion Ratio Lower Upper
98 100
100 65.6 52.1 78.1



#62
4-Bromofluorobenzene
Concen: 43.780 ug/l
RT: 12.626 min Scan# 1820
Delta R.T. -0.000 min
Lab File: VD071663.D
Acq: 10 Jan 2022 17:55

Tgt Ion: 95 Resp: 238395
Ion Ratio Lower Upper
95 100
174 75.8 0.0 150.0
176 71.9 0.0 146.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.638 min Scan# 117
Delta R.T. 0.001 min
Lab File: VD071663.D
Acq: 10 Jan 2022 17:55

Instrument :
MSVOA_D
ClientSampleId :
OK-03-011022

Tgt Ion:117 Resp: 566431
Ion Ratio Lower Upper
117 100
82 58.4 46.9 70.3
119 31.4 26.6 39.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.567 min Scan# 1980
Delta R.T. 0.000 min
Lab File: VD071663.D
Acq: 10 Jan 2022 17:55

Tgt Ion:152 Resp: 224966
Ion Ratio Lower Upper
152 100
115 62.9 31.5 94.5
150 160.3 0.0 348.0

