



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011124\
Data File : VD078142.D
Acq On : 11 Jan 2024 08:48
Operator : RP/MD
Sample : BFB
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D0108245.M
Title : SW846 8260
Last Update : Tue Jan 09 01:29:14 2024



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.4	16752	PASS
75	95	30	60	55.0	42973	PASS
95	95	100	100	100.0	78112	PASS
96	95	5	9	6.6	5139	PASS
173	174	0.00	2	1.3	871	PASS
174	95	50	100	85.1	66507	PASS
175	174	5	9	8.5	5639	PASS
176	174	95	101	98.6	65560	PASS
177	176	5	9	6.7	4366	PASS