

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD011819\
 Data File : VD060760.D
 Acq On : 18 Jan 2019 14:01
 Operator : VA/AP
 Sample : I.BLANK
 Misc : 5.00g/5ml/MSVOA D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 I.BLANK

Quant Time: Jan 19 03:26:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D011819S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 17 12:34:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.80	168	1578167	50.00	ug/l	0.02
34) 1,4-Difluorobenzene	7.98	114	2137845	50.00	ug/l	0.01
63) Chlorobenzene-d5	12.16	117	1723134	50.00	ug/l	0.02
72) 1,4-Dichlorobenzene-d4	14.32	152	867339	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.22	65	897329	82.28	ug/l	0.02
Spiked Amount 50.000			Recovery	=	164.56%	
35) Dibromofluoromethane	6.69	113	869800	55.59	ug/l	0.02
Spiked Amount 50.000			Recovery	=	111.18%	
50) Toluene-d8	10.18	98	2466314	52.83	ug/l	0.01
Spiked Amount 50.000			Recovery	=	105.66%	
62) 4-Bromofluorobenzene	13.35	95	952773	58.27	ug/l	0.02
Spiked Amount 50.000			Recovery	=	116.54%	
Target Compounds						
20) Methylene Chloride	3.69	84	36253	1.833	ug/l #	85
95) Naphthalene	15.93	128	42071	1.853	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD011819\
Data File : VD060760.D
Acq On : 18 Jan 2019 14:01
Operator : VA/AP
Sample : I.BLANK
Misc : 5.00µl/5ml/MSVOA D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
I.BLANK

Quant Time: Jan 19 03:26:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D011819S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 17 12:34:41 2019
Response via : Initial Calibration

