

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011922\
 Data File : VD071737.D
 Acq On : 19 Jan 2022 11:07
 Operator : VA/SY
 Sample : VD0119SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0119SBL01

Quant Time: Jan 20 04:56:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 07 15:39:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	258891	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	462498	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	427545	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	184723	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	194610	58.421	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.840%
35) Dibromofluoromethane	7.920	113	159481	52.177	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.360%
50) Toluene-d8	10.338	98	579378	50.638	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.280%
62) 4-Bromofluorobenzene	12.626	95	182295	45.215	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	90.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011922\
Data File : VD071737.D
Acq On : 19 Jan 2022 11:07
Operator : VA/SY
Sample : VD0119SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0119SBL01

Quant Time: Jan 20 04:56:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 07 15:39:14 2022
Response via : Initial Calibration

