

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011922\
 Data File : VD071752.D
 Acq On : 19 Jan 2022 19:01
 Operator : VA/SY
 Sample : VIBLK
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Jan 20 05:00:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 07 15:39:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	314910	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	591409	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	543463	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	226448	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	216987	53.551	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 107.100%	
35) Dibromofluoromethane	7.914	113	197658	50.572	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 101.140%	
50) Toluene-d8	10.332	98	767326	52.446	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 104.900%	
62) 4-Bromofluorobenzene	12.626	95	228084	44.241	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 88.480%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011922\
Data File : VD071752.D
Acq On : 19 Jan 2022 19:01
Operator : VA/SY
Sample : VIBLK
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Jan 20 05:00:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 07 15:39:14 2022
Response via : Initial Calibration

