

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011923\
 Data File : VD075133.D
 Acq On : 19 Jan 2023 10:50
 Operator : KP/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 20 00:35:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 19 01:32:27 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	108	0.00
2 T	Dichlorodifluoromethane	0.435	0.389	10.6	115	0.00
3 P	Chloromethane	0.646	0.590	8.7	105	0.00
4 C	Vinyl Chloride	0.693	0.645	6.9#	106	0.00
5 T	Bromomethane	0.471	0.433	8.1	104	0.00
6 T	Chloroethane	0.481	0.464	3.5	106	0.00
7 T	Trichlorofluoromethane	0.881	0.867	1.6	110	0.00
8 T	Diethyl Ether	0.284	0.296	-4.2	107	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.544	0.559	-2.8	117	0.00
10 T	Methyl Iodide	0.732	0.715	2.3	97	0.00
11 T	Tert butyl alcohol	0.036	0.032	11.1	104	0.00
12 CM	1,1-Dichloroethene	0.573	0.574	-0.2#	110	0.00
13 T	Acrolein	0.035	0.028	20.0	104	0.00
14 T	Allyl chloride	0.929	0.942	-1.4	106	0.00
15 T	Acrylonitrile	0.120	0.128	-6.7	107	0.00
16 T	Acetone	0.135	0.158	-17.0	106	0.00
17 T	Carbon Disulfide	1.857	1.862	-0.3	110	0.00
18 T	Methyl Acetate	0.366	0.369	-0.8	103	0.00
19 T	Methyl tert-butyl Ether	1.196	1.256	-5.0	105	0.00
20 T	Methylene Chloride	0.849	0.671	21.0	91	0.00
21 T	trans-1,2-Dichloroethene	0.637	0.646	-1.4	106	0.00
22 T	Diisopropyl ether	1.729	1.786	-3.3	106	0.00
23 T	Vinyl Acetate	0.762	0.809	-6.2	111	0.00
24 P	1,1-Dichloroethane	1.087	1.104	-1.6	107	0.00
25 T	2-Butanone	0.163	0.181	-11.0	107	0.00
26 T	2,2-Dichloropropane	0.969	0.948	2.2	107	0.00
27 T	cis-1,2-Dichloroethene	0.695	0.716	-3.0	107	0.00
28 T	Bromochloromethane	0.297	0.288	3.0	105	0.00
29 T	Tetrahydrofuran	0.093	0.097	-4.3	104	0.00
30 C	Chloroform	1.084	1.100	-1.5#	106	0.00
31 T	Cyclohexane	1.023	0.984	3.8	114	0.00
32 T	1,1,1-Trichloroethane	0.934	0.939	-0.5	108	0.00
33 S	1,2-Dichloroethane-d4	0.510	0.534	-4.7	107	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	109	0.00
35 S	Dibromofluoromethane	0.302	0.314	-4.0	106	0.00
36 T	1,1-Dichloropropene	0.485	0.480	1.0	110	0.00
37 T	Ethyl Acetate	0.196	0.207	-5.6	111	0.00
38 T	Carbon Tetrachloride	0.409	0.419	-2.4	113	0.00
39 T	Methylcyclohexane	0.598	0.600	-0.3	115	0.00
40 TM	Benzene	1.384	1.405	-1.5	107	0.00
41 T	Methacrylonitrile	0.118	0.129	-9.3	123	0.00
42 TM	1,2-Dichloroethane	0.364	0.366	-0.5	104	0.00
43 T	Isopropyl Acetate	0.381	0.388	-1.8	106	0.00
44 TM	Trichloroethene	0.389	0.384	1.3	105	0.00
45 C	1,2-Dichloropropane	0.341	0.346	-1.5#	106	0.00
46 T	Dibromomethane	0.180	0.187	-3.9	106	0.00
47 T	Bromodichloromethane	0.446	0.451	-1.1	104	0.00
48 T	Methyl methacrylate	0.186	0.186	0.0	107	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011923\
 Data File : VD075133.D
 Acq On : 19 Jan 2023 10:50
 Operator : KP/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 20 00:35:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 19 01:32:27 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	111	0.00
50 S	Toluene-d8	1.190	1.220	-2.5	108	0.00
51 T	4-Methyl-2-Pentanone	0.183	0.191	-4.4	109	0.00
52 CM	Toluene	0.875	0.870	0.6#	104	0.00
53 T	t-1,3-Dichloropropene	0.422	0.426	-0.9	104	0.00
54 T	cis-1,3-Dichloropropene	0.516	0.535	-3.7	105	0.00
55 T	1,1,2-Trichloroethane	0.238	0.244	-2.5	104	0.00
56 T	Ethyl methacrylate	0.286	0.300	-4.9	105	0.00
57 T	1,3-Dichloropropane	0.409	0.421	-2.9	103	0.00
58 T	2-Chloroethyl Vinyl ether	0.098	0.102	-4.1	115	0.00
59 T	2-Hexanone	0.132	0.144	-9.1	109	0.00
60 T	Dibromochloromethane	0.287	0.290	-1.0	102	0.00
61 T	1,2-Dibromoethane	0.227	0.229	-0.9	104	0.00
62 S	4-Bromofluorobenzene	0.378	0.384	-1.6	106	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	107	0.00
64 T	Tetrachloroethene	0.357	0.353	1.1	108	0.00
65 PM	Chlorobenzene	1.025	1.016	0.9	103	0.00
66 T	1,1,1,2-Tetrachloroethane	0.345	0.352	-2.0	108	0.00
67 C	Ethyl Benzene	1.844	1.844	0.0	106	0.00
68 T	m/p-Xylenes	0.712	0.708	0.6	105	0.00
69 T	o-Xylene	0.668	0.667	0.1	102	0.00
70 T	Styrene	1.111	1.116	-0.5	103	0.00
71 P	Bromoform	0.178	0.192	-7.9	109	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
73 T	Isopropylbenzene	3.949	3.936	0.3	107	0.00
74 T	N-amyl acetate	0.865	0.853	1.4	104	0.00
75 P	1,1,2,2-Tetrachloroethane	0.629	0.639	-1.6	104	0.00
76 T	1,2,3-Trichloropropane	0.423	0.416	1.7	110	0.00
77 T	Bromobenzene	0.902	0.861	4.5	98	0.00
78 T	n-propylbenzene	4.853	4.809	0.9	107	0.00
79 T	2-Chlorotoluene	2.669	2.626	1.6	104	0.00
80 T	1,3,5-Trimethylbenzene	3.236	3.232	0.1	105	0.00
81 T	trans-1,4-Dichloro-2-butene	0.183	0.183	0.0	105	0.00
82 T	4-Chlorotoluene	2.744	2.671	2.7	103	0.00
83 T	tert-Butylbenzene	2.800	2.765	1.2	107	0.00
84 T	1,2,4-Trimethylbenzene	3.178	3.101	2.4	103	0.00
85 T	sec-Butylbenzene	4.253	4.122	3.1	105	0.00
86 T	p-Isopropyltoluene	3.482	3.407	2.2	105	0.00
87 T	1,3-Dichlorobenzene	1.749	1.700	2.8	102	0.00
88 T	1,4-Dichlorobenzene	1.720	1.708	0.7	104	0.00
89 T	n-Butylbenzene	3.393	3.345	1.4	107	0.00
90 T	Hexachloroethane	0.615	0.600	2.4	107	0.00
91 T	1,2-Dichlorobenzene	1.496	1.469	1.8	101	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.096	0.100	-4.2	102	0.00
93 T	1,2,4-Trichlorobenzene	0.982	0.928	5.5	98	0.00
94 T	Hexachlorobutadiene	0.481	0.459	4.6	102	0.00
95 T	Naphthalene	1.956	1.834	6.2	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011923\
Data File : VD075133.D
Acq On : 19 Jan 2023 10:50
Operator : KP/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Jan 20 00:35:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011823S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 19 01:32:27 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.838	0.795	5.1	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 5