

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012023\
 Data File : VD075160.D
 Acq On : 20 Jan 2023 18:22
 Operator : KP/SY
 Sample : 01233-07
 Misc : 5.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 B-P37B

Quant Time: Jan 23 00:33:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 19 01:32:27 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	694	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	1678	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	1420	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	166	50.000	ug/l	# 0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.240	65	451	63.685	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	127.360%
35) Dibromofluoromethane	7.810	113	331	32.644	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	65.280%
50) Toluene-d8	10.269	98	2126	53.252	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.500%
62) 4-Bromofluorobenzene	12.581	95	509	40.172	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	80.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012023\
Data File : VD075160.D
Acq On : 20 Jan 2023 18:22
Operator : KP/SY
Sample : 01233-07
Misc : 5.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
B-P37B

Quant Time: Jan 23 00:33:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011823S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 19 01:32:27 2023
Response via : Initial Calibration

