

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012023\
 Data File : VD075165.D
 Acq On : 20 Jan 2023 20:41
 Operator : KP/SY
 Sample : 01258-01
 Misc : 3.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WASTE-VOA-2

Quant Time: Jan 23 00:35:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 19 01:32:27 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	24723	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	48713	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.587	117	48245	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	20922	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	19532	77.422	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	154.840%
35) Dibromofluoromethane	7.810	113	16653	56.574	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.140%
50) Toluene-d8	10.275	98	60592	52.280	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.560%
62) 4-Bromofluorobenzene	12.575	95	21856	59.419	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	118.840%
Target Compounds						
27) cis-1,2-Dichloroethene	7.081	96	20263	58.986	ug/l	91
44) Trichloroethene	9.034	130	12721	33.541	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012023\
Data File : VD075165.D
Acq On : 20 Jan 2023 20:41
Operator : KP/SY
Sample : 01258-01
Misc : 3.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WASTE-VOA-2

Quant Time: Jan 23 00:35:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011823S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 19 01:32:27 2023
Response via : Initial Calibration

