

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012023\
 Data File : VD075166.D
 Acq On : 20 Jan 2023 21:08
 Operator : KP/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 23 00:35:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 19 01:32:27 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	0.435	0.385	11.5	101	0.00
3 P	Chloromethane	0.646	0.593	8.2	94	0.00
4 C	Vinyl Chloride	0.693	0.586	15.4#	86	0.00
5 T	Bromomethane	0.471	0.407	13.6	87	0.00
6 T	Chloroethane	0.481	0.433	10.0	88	0.00
7 T	Trichlorofluoromethane	0.881	0.887	-0.7	100	0.00
8 T	Diethyl Ether	0.284	0.295	-3.9	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.544	0.551	-1.3	102	0.00
10 T	Methyl Iodide	0.732	0.735	-0.4	88	0.00
11 T	Tert butyl alcohol	0.036	0.030	16.7	87	0.01
12 CM	1,1-Dichloroethene	0.573	0.577	-0.7#	98	0.00
13 T	Acrolein	0.035	0.025	28.6#	80	0.00
14 T	Allyl chloride	0.929	0.995	-7.1	99	0.00
15 T	Acrylonitrile	0.120	0.134	-11.7	99	0.00
16 T	Acetone	0.135	0.128	5.2	77	0.00
17 T	Carbon Disulfide	1.857	1.866	-0.5	98	0.00
18 T	Methyl Acetate	0.366	0.400	-9.3	99	0.00
19 T	Methyl tert-butyl Ether	1.196	1.305	-9.1	97	0.00
20 T	Methylene Chloride	0.849	0.774	8.8	94	0.00
21 T	trans-1,2-Dichloroethene	0.637	0.670	-5.2	98	0.00
22 T	Diisopropyl ether	1.729	1.920	-11.0	101	0.00
23 T	Vinyl Acetate	0.762	0.758	0.5	93	0.00
24 P	1,1-Dichloroethane	1.087	1.173	-7.9	101	0.00
25 T	2-Butanone	0.163	0.168	-3.1	89	0.00
26 T	2,2-Dichloropropane	0.969	0.932	3.8	94	0.00
27 T	cis-1,2-Dichloroethene	0.695	0.752	-8.2	100	0.00
28 T	Bromochloromethane	0.297	0.304	-2.4	98	0.00
29 T	Tetrahydrofuran	0.093	0.101	-8.6	96	0.00
30 C	Chloroform	1.084	1.172	-8.1#	100	0.00
31 T	Cyclohexane	1.023	1.040	-1.7	107	0.00
32 T	1,1,1-Trichloroethane	0.934	0.996	-6.6	102	0.00
33 S	1,2-Dichloroethane-d4	0.510	0.582	-14.1	104	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	101	0.00
35 S	Dibromofluoromethane	0.302	0.315	-4.3	99	0.00
36 T	1,1-Dichloropropene	0.485	0.488	-0.6	104	0.00
37 T	Ethyl Acetate	0.196	0.202	-3.1	101	0.00
38 T	Carbon Tetrachloride	0.409	0.404	1.2	101	0.00
39 T	Methylcyclohexane	0.598	0.591	1.2	105	0.00
40 TM	Benzene	1.384	1.427	-3.1	101	0.00
41 T	Methacrylonitrile	0.118	0.111	5.9	98	0.00
42 TM	1,2-Dichloroethane	0.364	0.381	-4.7	100	0.00
43 T	Isopropyl Acetate	0.381	0.389	-2.1	98	0.00
44 TM	Trichloroethene	0.389	0.382	1.8	96	0.00
45 C	1,2-Dichloropropane	0.341	0.358	-5.0#	102	0.00
46 T	Dibromomethane	0.180	0.185	-2.8	98	0.00
47 T	Bromodichloromethane	0.446	0.465	-4.3	100	0.00
48 T	Methyl methacrylate	0.186	0.198	-6.5	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012023\
 Data File : VD075166.D
 Acq On : 20 Jan 2023 21:08
 Operator : KP/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 23 00:35:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 19 01:32:27 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	98	0.00
50 S	Toluene-d8	1.190	1.238	-4.0	102	0.00
51 T	4-Methyl-2-Pentanone	0.183	0.193	-5.5	102	0.00
52 CM	Toluene	0.875	0.900	-2.9#	99	0.00
53 T	t-1,3-Dichloropropene	0.422	0.429	-1.7	97	0.00
54 T	cis-1,3-Dichloropropene	0.516	0.531	-2.9	97	0.00
55 T	1,1,2-Trichloroethane	0.238	0.250	-5.0	99	0.00
56 T	Ethyl methacrylate	0.286	0.308	-7.7	100	0.00
57 T	1,3-Dichloropropane	0.409	0.429	-4.9	97	0.00
58 T	2-Chloroethyl Vinyl ether	0.098	0.099	-1.0	103	0.00
59 T	2-Hexanone	0.132	0.130	1.5	92	0.00
60 T	Dibromochloromethane	0.287	0.296	-3.1	97	0.00
61 T	1,2-Dibromoethane	0.227	0.238	-4.8	100	0.00
62 S	4-Bromofluorobenzene	0.378	0.400	-5.8	102	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	99	0.00
64 T	Tetrachloroethene	0.357	0.354	0.8	100	0.00
65 PM	Chlorobenzene	1.025	1.049	-2.3	99	0.00
66 T	1,1,1,2-Tetrachloroethane	0.345	0.352	-2.0	100	0.00
67 C	Ethyl Benzene	1.844	1.884	-2.2#	100	0.00
68 T	m/p-Xylenes	0.712	0.734	-3.1	102	0.00
69 T	o-Xylene	0.668	0.683	-2.2	98	0.00
70 T	Styrene	1.111	1.157	-4.1	100	0.00
71 P	Bromoform	0.178	0.186	-4.5	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
73 T	Isopropylbenzene	3.949	3.976	-0.7	102	0.00
74 T	N-amyl acetate	0.865	0.835	3.5	95	0.00
75 P	1,1,2,2-Tetrachloroethane	0.629	0.647	-2.9	99	0.00
76 T	1,2,3-Trichloropropane	0.423	0.492	-16.3	123	0.00
77 T	Bromobenzene	0.902	0.875	3.0	94	0.00
78 T	n-propylbenzene	4.853	4.867	-0.3	102	0.00
79 T	2-Chlorotoluene	2.669	2.677	-0.3	100	0.00
80 T	1,3,5-Trimethylbenzene	3.236	3.279	-1.3	100	0.00
81 T	trans-1,4-Dichloro-2-butene	0.183	0.174	4.9	94	0.00
82 T	4-Chlorotoluene	2.744	2.697	1.7	98	0.00
83 T	tert-Butylbenzene	2.800	2.746	1.9	100	0.00
84 T	1,2,4-Trimethylbenzene	3.178	3.225	-1.5	101	0.00
85 T	sec-Butylbenzene	4.253	4.248	0.1	102	0.00
86 T	p-Isopropyltoluene	3.482	3.441	1.2	100	0.00
87 T	1,3-Dichlorobenzene	1.749	1.702	2.7	96	0.00
88 T	1,4-Dichlorobenzene	1.720	1.703	1.0	98	0.00
89 T	n-Butylbenzene	3.393	3.373	0.6	101	0.00
90 T	Hexachloroethane	0.615	0.569	7.5	95	0.00
91 T	1,2-Dichlorobenzene	1.496	1.498	-0.1	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.096	0.097	-1.0	93	0.00
93 T	1,2,4-Trichlorobenzene	0.982	0.917	6.6	91	0.00
94 T	Hexachlorobutadiene	0.481	0.451	6.2	94	0.00
95 T	Naphthalene	1.956	2.011	-2.8	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012023\
Data File : VD075166.D
Acq On : 20 Jan 2023 21:08
Operator : KP/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Jan 23 00:35:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011823S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 19 01:32:27 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.838	0.774	7.6	90	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6