

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012224\
 Data File : VD078200.D
 Acq On : 22 Jan 2024 19:32
 Operator : RP/MD
 Sample : BLK
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BLK

Quant Time: Jan 23 05:50:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012224S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 23 05:43:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	153231	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.781	114	311518	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.587	117	299869	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	124145	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	68643	39.241	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	78.480%
35) Dibromofluoromethane	7.810	113	88753	43.290	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	86.580%
50) Toluene-d8	10.275	98	385261	47.457	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.920%
62) 4-Bromofluorobenzene	12.575	95	108619	46.137	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	92.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012224\
Data File : VD078200.D
Acq On : 22 Jan 2024 19:32
Operator : RP/MD
Sample : BLK
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BLK

Quant Time: Jan 23 05:50:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012224S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 23 05:43:07 2024
Response via : Initial Calibration

