

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD012420\
 Data File : VD064940.D
 Acq On : 24 Jan 2020 13:48
 Operator : VA/SY
 Sample : L1243-01
 Misc : 6.20G/5.00ml/MSVOA D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CTY-SB-06-1.5-2.5

Quant Time: Jan 25 03:01:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D012320S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 23 13:45:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	495971	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	793427	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	710291	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	295701	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.34	65	219886	46.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.44%	
35) Dibromofluoromethane	7.92	113	231352	46.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.80%	
50) Toluene-d8	10.34	98	942196	48.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.92%	
62) 4-Bromofluorobenzene	12.64	95	289835	49.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.44%	
Target Compounds						
16) Acetone	4.17	43	20205	28.279	ug/l	98
17) Carbon Disulfide	4.40	76	44719	3.147	ug/l	97
44) Trichloroethene	9.11	130	8348	1.416	ug/l	79
64) Tetrachloroethene	10.88	164	7937	1.546	ug/l	93
95) Naphthalene	15.41	128	11222	1.248	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD012420\
Data File : VD064940.D
Acq On : 24 Jan 2020 13:48
Operator : VA/SY
Sample : L1243-01
Misc : 6.20G/5.00ml/MSVOA D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CTY-SB-06-1.5-2.5

Quant Time: Jan 25 03:01:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D012320S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 23 13:45:59 2020
Response via : Initial Calibration

