

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD012420\
 Data File : VD064944.D
 Acq On : 24 Jan 2020 15:43
 Operator : VA/SY
 Sample : L1243-05
 Misc : 6.20G/5.00ml/MSVOA D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CTY-SB-03-1-2

Quant Time: Jan 25 03:07:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D012320S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 23 13:45:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	454911	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	731023	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	669295	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	279590	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.34	65	209015	47.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.80%	
35) Dibromofluoromethane	7.92	113	220292	47.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.90%	
50) Toluene-d8	10.34	98	883908	49.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.70%	
62) 4-Bromofluorobenzene	12.64	95	263530	48.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.14%	
Target Compounds						
16) Acetone	4.17	43	19820	29.294	ug/l #	87
17) Carbon Disulfide	4.40	76	19339	1.484	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD012420\
Data File : VD064944.D
Acq On : 24 Jan 2020 15:43
Operator : VA/SY
Sample : L1243-05
Misc : 6.20G/5.00ml/MSVOA D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CTY-SB-03-1-2

Quant Time: Jan 25 03:07:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D012320S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 23 13:45:59 2020
Response via : Initial Calibration

