

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD012420\
 Data File : VD064947.D
 Acq On : 24 Jan 2020 17:09
 Operator : VA/SY
 Sample : L1243-10
 Misc : 6.77G/5.00ml/MSVOA D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CTY-SB-03-5-6

Quant Time: Jan 25 03:14:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D012320S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 23 13:45:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	453097	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	718638	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	633506	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	240168	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.34	65	201879	46.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.90%	
35) Dibromofluoromethane	7.92	113	218435	48.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.74%	
50) Toluene-d8	10.34	98	861372	49.42	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.84%	
62) 4-Bromofluorobenzene	12.64	95	243187	45.59	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.18%	
Target Compounds						
16) Acetone	4.17	43	18839	28.580	ug/l	94
17) Carbon Disulfide	4.41	76	27334	2.106	ug/l #	94
39) Methylcyclohexane	9.36	83	39731	4.904	ug/l	95
40) Benzene	8.36	78	68043	3.565	ug/l	97
68) m/p-Xylenes	11.86	106	8853	1.007	ug/l	91
84) 1,2,4-Trimethylbenzene	13.27	105	18809	1.380	ug/l	99
95) Naphthalene	15.41	128	442382	60.562	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD012420\
Data File : VD064947.D
Acq On : 24 Jan 2020 17:09
Operator : VA/SY
Sample : L1243-10
Misc : 6.77G/5.00ml/MSVOA D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CTY-SB-03-5-6

Quant Time: Jan 25 03:14:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D012320S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 23 13:45:59 2020
Response via : Initial Calibration

