

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD012420\
 Data File : VD064952.D
 Acq On : 24 Jan 2020 19:33
 Operator : VA/SY
 Sample : L1242-08
 Misc : 5.08G/5.00ml/MSVOA D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 20200061-COMPOSITE-B7

Quant Time: Jan 25 03:30:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D012320S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 23 13:45:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	429134	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	700638	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	641210	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	262358	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.34	65	211831	51.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.92%	
35) Dibromofluoromethane	7.92	113	221062	50.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.42%	
50) Toluene-d8	10.34	98	834428	49.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.20%	
62) 4-Bromofluorobenzene	12.64	95	254048	48.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.70%	
Target Compounds						
16) Acetone	4.17	43	71924	75.239	ug/l	97
17) Carbon Disulfide	4.41	76	25382	2.065	ug/l #	92
25) 2-Butanone	7.22	43	26190	23.550	ug/l	95
95) Naphthalene	15.41	128	107184	13.432	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD012420\
Data File : VD064952.D
Acq On : 24 Jan 2020 19:33
Operator : VA/SY
Sample : L1242-08
Misc : 5.08G/5.00ml/MSVOA D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
20200061-COMPOSITE-B7

Quant Time: Jan 25 03:30:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D012320S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 23 13:45:59 2020
Response via : Initial Calibration

