

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012523\
 Data File : VD075198.D
 Acq On : 25 Jan 2023 12:17
 Operator : KP/SY
 Sample : 01268-01RE
 Misc : 5.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TR-06-012323RE

Quant Time: Jan 25 15:56:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 19 01:32:27 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	32280	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	62169	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.587	117	62908	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	28025	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	23156	70.299	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	140.600%
35) Dibromofluoromethane	7.804	113	21212	56.465	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.920%
50) Toluene-d8	10.269	98	77530	52.415	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.840%
62) 4-Bromofluorobenzene	12.575	95	26045	55.481	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	110.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012523\
Data File : VD075198.D
Acq On : 25 Jan 2023 12:17
Operator : KP/SY
Sample : 01268-01RE
Misc : 5.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TR-06-012323RE

Quant Time: Jan 25 15:56:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011823S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 19 01:32:27 2023
Response via : Initial Calibration

