

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012523\
 Data File : VD075202.D
 Acq On : 25 Jan 2023 14:08
 Operator : KP/SY
 Sample : 01251-05
 Misc : 3.98G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TOT-VOC-1

Quant Time: Jan 25 15:57:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 19 01:32:27 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	45753	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	88017	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	87856	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	40405	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	32686	70.010	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 140.020%		
35) Dibromofluoromethane	7.810	113	29512	55.488	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 110.980%		
50) Toluene-d8	10.269	98	111717	53.348	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 106.700%		
62) 4-Bromofluorobenzene	12.575	95	39189	58.965	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 117.940%		
Target Compounds						
16) Acetone	4.022	43	9752	79.197	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012523\
Data File : VD075202.D
Acq On : 25 Jan 2023 14:08
Operator : KP/SY
Sample : 01251-05
Misc : 3.98G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TOT-VOC-1

Quant Time: Jan 25 15:57:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011823S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 19 01:32:27 2023
Response via : Initial Calibration

